



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85656

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/02/2017	17:30
2	CAL2	1	Water			7.00	02/02/2017	17:31
3	CAL3	1	Water			10.02	02/02/2017	17:32
4	ICV	1	Water			7.03	02/02/2017	17:33
5	CCV1	1	Water			2.02	02/02/2017	17:34
6	I1687-03	1	Solid	20.02	20	7.58	02/02/2017	17:35
7	I1687-03DUP	1	Solid	20.04	20	7.59	02/02/2017	17:36
8	I1687-04	1	Solid	20.01	20	8.25	02/02/2017	17:37
9	CCV2	1	Water			12.02	02/02/2017	17:38

1385656

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1687 WorkList ID : 95293 Date : 2/2/2017 5:28:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1687-03	Corrosivity	Cool 4 deg C	ENTA04	G32	WC-B6-B7-001	02/01/2017	9045C
	Solid	I1687-04	Corrosivity	Cool 4 deg C	ENTA04	G32	WC-B25-B30-001	02/01/2017	9045C

Date/Time 02/02/17 5:28 PM
Received by: n
Relinquished by: [Signature]

Date/Time 02/02/17 6:50 PM
Received by: [Signature]
Relinquished by: 31