


7470A

LB85667
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36939



A= 0.0000e+000

B= 2.7009e-004

C= -3.1023e-002

Rho= 0.9999178

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	73	0.000	73				
0.2	0.200	0.162	-0.038	715	0.0 %	715				
2.5	2.500	2.501	0.001	9374	0.0 %	9374				
5.0	5.000	5.094	0.094	18976	0.0 %	18976				
7.5	7.500	7.508	0.008	27913	0.0 %	27913				
10.0	10.000	9.946	-0.054	36939	0.0 %	36939				

LB85667 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Feb 2017 11:54:27

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	73	-	0.00007470A	PPB	03 Feb 2017 12:08:04	S	Std
0.2 - 1	50.2	715	-	0.20007470A	PPB	03 Feb 2017 12:10:12	S	Std
2.5 - 1	52.5	9374	-	2.50007470A	PPB	03 Feb 2017 12:12:21	S	Std
5.0 - 1	55	18976	-	5.00007470A	PPB	03 Feb 2017 12:14:29	S	Std
7.5 - 1	57.5	27913	-	7.50007470A	PPB	03 Feb 2017 12:16:37	S	Std
10.0 - 1	59	36939	-	10.00007470A	PPB	03 Feb 2017 12:18:44	S	Std
ICV17 - 1	ICV17	15134	4.0566	-7470A	PPB	03 Feb 2017 12:21:29	U	SMPL
ICB17 - 1	ICB17	17	-0.0264	-7470A	PPB	03 Feb 2017 12:23:35	U	SMPL
CCV58 - 1	CCV58	19113	5.1313	-7470A	PPB	03 Feb 2017 12:25:42	U	SMPL
CCB58 - 1	CCB58	22	-0.0251	-7470A	PPB	03 Feb 2017 12:27:49	U	SMPL
CRA - 1	CRA	838	0.1953	-7470A	PPB	03 Feb 2017 12:32:17	U	SMPL
HighStd - 1	HighStd	37072	9.9819	-7470A	PPB	03 Feb 2017 12:34:23	U	SMPL
ChkStd - 1	ChkStd	25537	6.8664	-7470A	PPB	03 Feb 2017 12:36:30	U	SMPL
PB96690BL - 1	PBW	-13	-0.0345	-7470A	PPB	03 Feb 2017 12:38:37	U	SMPL
PB96690BS - 1	LCSW	12941	3.4643	-7470A	PPB	03 Feb 2017 12:40:44	U	SMPL
PB96690TB - 1	PB96659TB	1	-0.0308	-7470A	PPB	03 Feb 2017 12:42:52	U	SMPL
I1658-02 - 1	SP-COMP-1A-0-2FT	21	-0.0254	-7470A	PPB	03 Feb 2017 12:45:00	U	SMPL
I1658-07 - 1	SP-COMP-2A-0-2FT	22	-0.0251	-7470A	PPB	03 Feb 2017 12:47:07	U	SMPL
I1658-12 - 1	SP-COMP-02A-0-2FT	46	-0.0186	-7470A	PPB	03 Feb 2017 12:49:16	U	SMPL
I1658-17 - 1	SP-COMP-02B-0-2FT	40	-0.0202	-7470A	PPB	03 Feb 2017 12:51:23	U	SMPL
CCV59 - 1	CCV59	19054	5.1153	-7470A	PPB	03 Feb 2017 12:53:30	U	SMPL
CCB59 - 1	CCB59	12	-0.0278	-7470A	PPB	03 Feb 2017 12:55:37	U	SMPL
I1659-07 - 1	CENTER-COMP	29	-0.0232	-7470A	PPB	03 Feb 2017 12:57:44	U	SMPL
I1661-03 - 1	WC-B32-B33-001	109	-0.0016	-7470A	PPB	03 Feb 2017 12:59:51	U	SMPL
I1661-04 - 1	WC-B28-B36-001	45	-0.0189	-7470A	PPB	03 Feb 2017 13:01:58	U	SMPL
I1664-02 - 1	ROLL-OFF-275032-COMP-0-2FT	12	-0.0278	-7470A	PPB	03 Feb 2017 13:04:05	U	SMPL
I1665-02 - 1	ROLL-OFF-2532-COMP-0-2FT	8	-0.0289	-7470A	PPB	03 Feb 2017 13:06:12	U	SMPL
I1665-02DUP - 1	ROLL-OFF-2532-COMP-0-2FTDUP	25	-0.0243	-7470A	PPB	03 Feb 2017 13:08:20	U	SMPL
I1665-02MS - 1	ROLL-OFF-2532-COMP-0-2FTMS	12851	3.4400	-7470A	PPB	03 Feb 2017 13:10:28	U	SMPL
I1665-02MSD - 1	ROLL-OFF-2532-COMP-0-2FTMSD	12883	3.4486	-7470A	PPB	03 Feb 2017 13:12:36	U	SMPL
I1665-02LX5 - 1		31	-0.0226	-7470A	PPB	03 Feb 2017 13:14:45	U	SMPL
I1665-02A - 1		15152	4.0614	-7470A	PPB	03 Feb 2017 13:16:52	U	SMPL
CCV60 - 1	CCV60	19111	5.1307	-7470A	PPB	03 Feb 2017 13:18:59	U	SMPL
CCB60 - 1	CCB60	21	-0.0254	-7470A	PPB	03 Feb 2017 13:21:06	U	SMPL