



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB85668

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	02/03/2017	11:00
2	I1272-03		1	85.0	185.0	02/03/2017	11:30
3	I1272-03DUP		1	85.0	185.0	02/03/2017	12:00

Result = (Celsius * 1.8) + 32