


ISM02.3

LB85678
INSTRUMENT ID: CV4

Linear

 μ Abs.:

4382E



A= 0.0000e+000

B= 2.2806e-004

C= -1.5641e-002

Rho= 0.9999929

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.012	-0.012	16	0.000	16					-
0.2	0.200	0.193	-0.007	914	0.0 %	914					-4
2.5	2.500	2.511	0.011	11077	0.0 %	11077					0
5.0	5.000	5.016	0.016	22061	0.0 %	22061					0
7.5	7.500	7.513	0.013	33013	0.0 %	33013					0
10.0	10.000	9.979	-0.021	43826	0.0 %	43826					0

AB
2/3/17

LB85678
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	16	-		0 ISM02.3	1	2/3/2017 14:25	Std
	0.2 S0.2	914	-		0.2 ISM02.3	1	2/3/2017 14:27	Std
	2.5 S2.5	11077	-		2.5 ISM02.3	1	2/3/2017 14:29	Std
	5 S5	22061	-		5 ISM02.3	1	2/3/2017 14:31	Std
	7.5 S7.5	33013	-		7.5 ISM02.3	1	2/3/2017 14:33	Std
	10 S10	43826	-		10 ISM02.3	1	2/3/2017 14:35	Std
ICV	ICV	17155	3.8968 -		ISM02.3	1	2/3/2017 14:38	SMPL
ICB	ICB	-9	-0.0177 -		ISM02.3	1	2/3/2017 14:40	SMPL
CCV061	CCV061	21672	4.927 -		ISM02.3	1	2/3/2017 14:42	SMPL
CCB061	CCB061	-17	-0.0195 -		ISM02.3	1	2/3/2017 14:44	SMPL
PB96680BL	PBW003	91	0.0051 -		ISM02.3	1	2/3/2017 14:47	SMPL
I1456-01	MJHAF7	149	0.0183 -		ISM02.3	1	2/3/2017 14:49	SMPL
I1456-02	MJHAF8	153	0.0193 -		ISM02.3	1	2/3/2017 14:51	SMPL
I1456-03	MJHAF9	95	0.006 -		ISM02.3	1	2/3/2017 14:53	SMPL
I1456-04	MJHAF9D	90	0.0049 -		ISM02.3	1	2/3/2017 14:55	SMPL
I1456-05	MJHAF9S	4305	0.9662 -		ISM02.3	1	2/3/2017 14:57	SMPL
I1456-06	MJHAG0	116	0.0108 -		ISM02.3	1	2/3/2017 14:59	SMPL
I1456-07	MJHAG1	81	0.0028 -		ISM02.3	1	2/3/2017 15:01	SMPL
I1456-08	MJHAG2	118	0.0113 -		ISM02.3	1	2/3/2017 15:04	SMPL
I1456-09	MJHAG3	75	0.0015 -		ISM02.3	1	2/3/2017 15:06	SMPL
PB96681BL	PBW003	107	0.0088 -		ISM02.3	1	2/3/2017 15:08	SMPL
I1458-01	MJHAF0	106	0.0085 -		ISM02.3	1	2/3/2017 15:10	SMPL
I1458-02	MJHAF1	96	0.0063 -		ISM02.3	1	2/3/2017 15:12	SMPL
I1458-03	MJHAF2	82	0.0031 -		ISM02.3	1	2/3/2017 15:14	SMPL
I1458-04	MJHAF2D	76	0.0017 -		ISM02.3	1	2/3/2017 15:16	SMPL
I1458-05	MJHAF2S	4084	0.9158 -		ISM02.3	1	2/3/2017 15:18	SMPL
I1458-06	MJHAF3	89	0.0047 -		ISM02.3	1	2/3/2017 15:21	SMPL
I1458-07	MJHAF4	80	0.0026 -		ISM02.3	1	2/3/2017 15:23	SMPL
I1458-08	MJHAF5	96	0.0063 -		ISM02.3	1	2/3/2017 15:25	SMPL
I1458-09	MJHAF6	79	0.0024 -		ISM02.3	1	2/3/2017 15:27	SMPL
CCV062	CCV062	21290	4.8398 -		ISM02.3	1	2/3/2017 15:29	SMPL
CCB062	CCB062	-4	-0.0166 -		ISM02.3	1	2/3/2017 15:31	SMPL
I1458-10	MJHAG4	84	0.0035 -		ISM02.3	1	2/3/2017 15:33	SMPL
I1458-11	MJHAG5	86	0.004 -		ISM02.3	1	2/3/2017 15:35	SMPL
I1458-12	MJHAG6	77	0.0019 -		ISM02.3	1	2/3/2017 15:38	SMPL
I1458-13	MJHAG7	87	0.0042 -		ISM02.3	1	2/3/2017 15:40	SMPL
I1458-14	MJHAG8	89	0.0047 -		ISM02.3	1	2/3/2017 15:42	SMPL
I1458-15	MJHAG9	104	0.0081 -		ISM02.3	1	2/3/2017 15:44	SMPL
I1458-16	MJHAH0	93	0.0056 -		ISM02.3	1	2/3/2017 15:46	SMPL
I1458-17	MJHAH1	109	0.0092 -		ISM02.3	1	2/3/2017 15:48	SMPL
PB96693BL	PBW003	94	0.0058 -		ISM02.3	1	2/3/2017 15:50	SMPL
I1497-01	MC0AA7	98	0.0067 -		ISM02.3	1	2/3/2017 15:52	SMPL
I1497-02	MC0AA7D	52	-0.0038 -		ISM02.3	1	2/3/2017 15:55	SMPL
I1497-03	MC0AA7S	4012	0.8994 -		ISM02.3	1	2/3/2017 15:57	SMPL
I1497-04	MC0AB0	46	-0.0051 -		ISM02.3	1	2/3/2017 15:59	SMPL
PB96694BL	PBW003	45	-0.0054 -		ISM02.3	1	2/3/2017 16:01	SMPL

LB85678
INSTRUMENT ID : CV1

I1533-01	ME60X6	31	-0.0086 -	ISM02.3	1	2/3/2017 16:03	SMPL
I1533-02	ME60X9	36	-0.0074 -	ISM02.3	1	2/3/2017 16:05	SMPL
I1533-03	ME60Y1	48	-0.0047 -	ISM02.3	1	2/3/2017 16:07	SMPL
I1533-04	ME60Z2	23	-0.0104 -	ISM02.3	1	2/3/2017 16:09	SMPL
I1533-05	ME6124	95	0.006 -	ISM02.3	1	2/3/2017 16:12	SMPL
I1533-06	ME6128	116	0.0108 -	ISM02.3	1	2/3/2017 16:14	SMPL
CCV063	CCV063	20957	4.7639 -	ISM02.3	1	2/3/2017 16:16	SMPL
CCB063	CCB063	-16	-0.0193 -	ISM02.3	1	2/3/2017 16:18	SMPL
I1533-07	ME6128D	92	0.0053 -	ISM02.3	1	2/3/2017 16:20	SMPL
I1533-08	ME6128S	4251	0.9539 -	ISM02.3	1	2/3/2017 16:22	SMPL
I1533-09	ME6137	68	-0.0001 -	ISM02.3	1	2/3/2017 16:24	SMPL
PB96695BL	PBW003	98	0.0067 -	ISM02.3	1	2/3/2017 16:26	SMPL
I1534-01	ME60X7	40	-0.0065 -	ISM02.3	1	2/3/2017 16:29	SMPL
I1534-02	ME60Y0	43	-0.0058 -	ISM02.3	1	2/3/2017 16:31	SMPL
I1534-03	ME60Y2	24	-0.0102 -	ISM02.3	1	2/3/2017 16:33	SMPL
I1534-04	ME60Z3	32	-0.0083 -	ISM02.3	1	2/3/2017 16:35	SMPL
I1534-05	ME6125	127	0.0133 -	ISM02.3	1	2/3/2017 16:37	SMPL
I1534-06	ME6129	118	0.0113 -	ISM02.3	1	2/3/2017 16:39	SMPL
I1534-07	ME6129D	116	0.0108 -	ISM02.3	1	2/3/2017 16:41	SMPL
I1534-08	ME6129S	4341	0.9744 -	ISM02.3	1	2/3/2017 16:43	SMPL
I1534-09	ME6138	106	0.0085 -	ISM02.3	1	2/3/2017 16:46	SMPL
CCV064	CCV064	21087	4.7935 -	ISM02.3	1	2/3/2017 16:48	SMPL
CCB064	CCB064	-8	-0.0175 -	ISM02.3	1	2/3/2017 16:50	SMPL