


245.1

LB 85682
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36649



A= 0.0000e+000

B= 2.7253e-004

C= -8.9482e-004

Rho= 0.9999931

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	23	0.000	23				
0.2	0.200	0.202	0.002	743	0.0 %	743				
2.5	2.500	2.487	-0.013	9130	0.0 %	9130				
5.0	5.000	4.992	-0.008	18321	0.0 %	18321				
7.5	7.500	7.527	0.027	27621	0.0 %	27621				
10.0	10.000	9.987	-0.013	36649	0.0 %	36649				

LB85682 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Feb 2017 17:06:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	23	-	0.0000	245.1	PPB	03 Feb 2017 17:12:59	S	Std
0.2 - 1	SO.2	743	-	0.2000	245.1	PPB	03 Feb 2017 17:15:07	S	Std
2.5 - 1	SO.5	9130	-	2.5000	245.1	PPB	03 Feb 2017 17:17:15	S	Std
5.0 - 1	SO.5	18321	-	5.0000	245.1	PPB	03 Feb 2017 17:19:23	S	Std
7.5 - 1	SO.5	27621	-	7.5000	245.1	PPB	03 Feb 2017 17:21:31	S	Std
10.0 - 1	SO	36649	-	10.0000	245.1	PPB	03 Feb 2017 17:23:38	S	Std
ICV18 - 1	ICV18	14984	4.0827	-245.1		PPB	03 Feb 2017 17:26:20	U	SMPL
ICB18 - 1	ICB18	-22	-0.0069	-245.1		PPB	03 Feb 2017 17:28:26	U	SMPL
CCV65 - 1	CCV65	18335	4.9959	-245.1		PPB	03 Feb 2017 17:30:33	U	SMPL
CCB65 - 1	CCB65	-3	-0.0017	-245.1		PPB	03 Feb 2017 17:32:39	U	SMPL
CRA - 1	CRA	730	0.1981	-245.1		PPB	03 Feb 2017 17:34:46	U	SMPL
HighStd - 1	HighStd	36193	9.8627	-245.1		PPB	03 Feb 2017 17:36:52	U	SMPL
ChkStd - 1	ChkStd	26425	7.2007	-245.1		PPB	03 Feb 2017 17:38:59	U	SMPL
PB96712BL - 1	PBW	-30	-0.0091	-245.1		PPB	03 Feb 2017 17:41:06	U	SMPL
PB96712BS - 1	LCSW	14113	3.8453	-245.1		PPB	03 Feb 2017 17:43:14	U	SMPL
I1307-08 - 1	PT-HG-WS	3146	0.8565	-245.1		PPB	03 Feb 2017 17:47:30	U	SMPL
CCV66 - 1	CCV66	18294	4.9847	-245.1		PPB	03 Feb 2017 17:53:54	U	SMPL
CCB66 - 1	CCB66	-10	-0.0036	-245.1		PPB	03 Feb 2017 17:56:01	U	SMPL