



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : AMANDEEP

Run Number: LB85684

Temp.: 22.3 °C

BalanceID: WC-SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/03/2017	16:35
2	CAL2	1	Water			7.01	02/03/2017	16:36
3	CAL3	1	Water			10.04	02/03/2017	16:37
4	ICV	1	Water			7.00	02/03/2017	16:38
5	CCV1	1	Water			2.02	02/03/2017	16:40
6	I1690-04	1	Solid	20.02	20	9.64	02/03/2017	16:41
7	I1690-04DUP	1	Solid	20.02	20	9.65	02/03/2017	16:42
8	I1697-06	1	Solid	20.04	20	7.45	02/03/2017	16:43
9	I1697-07	1	Solid	20.03	20	7.20	02/03/2017	16:44
10	I1697-08	1	Solid	20.01	20	5.66	02/03/2017	16:45
11	I1697-09	1	Solid	20.05	20	7.35	02/03/2017	16:46
12	I1697-10	1	Solid	20.04	20	7.75	02/03/2017	16:47
13	I1698-05	1	Solid	20.01	20	9.34	02/03/2017	16:48
14	I1698-06	1	Solid	20.02	20	7.35	02/03/2017	16:50
15	I1698-07	1	Solid	20.00	20	8.05	02/03/2017	16:51
16	CCV2	1	Water			12.02	02/03/2017	16:52
17	I1698-08	1	Solid	20.03	20	7.17	02/03/2017	16:53
18	CCV3	1	Water			12.01	02/03/2017	16:54

LB85684

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 11690 WorkList ID : 95322 Date : 2/3/2017 1:41:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/07/2017	Solid	I1690-04	Corrosivity	Cool 4 deg C	PSEG01	E42	MH-6-COMP	02/03/2017	9045C
	Solid	I1697-06	Corrosivity	Cool 4 deg C	ENTA04	G51	WC-B10-B17-001	02/03/2017	9045C
	Solid	I1697-07	Corrosivity	Cool 4 deg C	ENTA04	G51	WC-B26-B27-001	02/03/2017	9045C
	Solid	I1697-08	Corrosivity	Cool 4 deg C	ENTA04	G51	WC-B18-B19-001	02/03/2017	9045C
	Solid	I1697-09	Corrosivity	Cool 4 deg C	ENTA04	G51	WC-B23-B24-001	02/03/2017	9045C
	Solid	I1697-10	Corrosivity	Cool 4 deg C	ENTA04	G51	WC-B20-B22-001	02/03/2017	9045C
	Solid	I1698-05	Corrosivity	Cool 4 deg C	ENTA04	G41	WC-B8-B12-001	02/02/2017	9045C
	Solid	I1698-06	Corrosivity	Cool 4 deg C	ENTA04	G41	WC-B4-B9-001	02/02/2017	9045C
	Solid	I1698-07	Corrosivity	Cool 4 deg C	ENTA04	G41	WC-B13-B14-001	02/02/2017	9045C
	Solid	I1698-08	Corrosivity	Cool 4 deg C	ENTA04	G41	WC-B11-B16-001	02/02/2017	9045C

Date/Time 02/03/17 1:41:17
Received by: JLL
Relinquished by: JLL

Date/Time 02/03/17 6:35 PM
Received by: CS
Relinquished by: JLL