


7471A

LB85685
INSTRUMENT ID: CV2

Linear

 μ Abs.:

4454E



A= 0.0000e+000

B= 2.2388e-004

C= 5.1524e-003

Rho= 0.9998776

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	7	0.000	7				
0.2	0.200	0.193	-0.007	837	0.0 %	837				
2.5	2.500	2.546	0.046	11348	0.0 %	11348				
5.0	5.000	4.896	-0.104	21847	0.0 %	21847				
7.5	7.500	7.581	0.081	33839	0.0 %	33839				
10.0	10.000	9.978	-0.022	44546	0.0 %	44546				

LB85685 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 03 Feb 2017 19:01:44

Sample ID	Extended ID	µ Abs.	Conc.	Strd ConcMethod	Units	Date	Type	Type
0.0 - 1	50	7	-	0.00007471A	PPB	03 Feb 2017 19:03:00	S	Std
0.2 - 1	80.2	837	-	0.20007471A	PPB	03 Feb 2017 19:05:18	S	Std
2.5 - 1	52.5	11348	-	2.50007471A	PPB	03 Feb 2017 19:07:36	S	Std
5.0 - 1	55	21847	-	5.00007471A	PPB	03 Feb 2017 19:09:53	S	Std
7.5 - 1	57.5	33839	-	7.50007471A	PPB	03 Feb 2017 19:12:10	S	Std
10.0 - 1	510	44546	-	10.00007471A	PPB	03 Feb 2017 19:14:27	S	Std
ICV19 - 1	ICV19	16820	3.7708	-7471A	PPB	03 Feb 2017 19:17:14	U	SMPL
ICB19 - 1	ICB19	-5	0.0040	-7471A	PPB	03 Feb 2017 19:19:29	U	SMPL
CCV67 - 1	CCV67	21534	4.8261	-7471A	PPB	03 Feb 2017 19:21:45	U	SMPL
CCB67 - 1	CCB67	0	0.0052	-7471A	PPB	03 Feb 2017 19:24:00	U	SMPL
CRA - 1	CRA	833	0.1916	-7471A	PPB	03 Feb 2017 19:26:16	U	SMPL
HighStd - 1	HighStd	44316	9.9265	-7471A	PPB	03 Feb 2017 19:28:32	U	SMPL
ChkStd - 1	ChkStd	30546	6.8437	-7471A	PPB	03 Feb 2017 19:30:47	U	SMPL
PB96717BL - 1	PBS	-18	0.0011	-7471A	PPB	03 Feb 2017 19:33:03	U	SMPL
PB96717BS - 1	LCSS	15881	3.5605	-7471A	PPB	03 Feb 2017 19:35:20	U	SMPL
I1658-01 - 1	SP-COMP-1A-0-2FT	9981	2.2397	-7471A	PPB	03 Feb 2017 19:37:37	U	SMPL
I1658-06 - 1	SP-COMP-2A-0-2FT	26187	5.8678	-7471A	PPB	03 Feb 2017 19:39:54	U	SMPL
I1658-11 - 1	SP-COMP-02A-0-2FT	15669	3.5131	-7471A	PPB	03 Feb 2017 19:42:12	U	SMPL
I1658-16 - 1	SP-COMP-02B-0-2FT	21024	4.7119	-7471A	PPB	03 Feb 2017 19:44:29	U	SMPL
I1659-06 - 1	CENTER-COMP	11950	2.6805	-7471A	PPB	03 Feb 2017 19:46:45	U	SMPL
CCV68 - 1	CCV68	21618	4.8449	-7471A	PPB	03 Feb 2017 19:49:01	U	SMPL
CCB68 - 1	CCB68	0	0.0052	-7471A	PPB	03 Feb 2017 19:51:17	U	SMPL
I1664-01 - 1	ROLL-OFF-275032-COMP-0-2FT	1848	0.4189	-7471A	PPB	03 Feb 2017 19:53:33	U	SMPL
I1665-01 - 1	ROLL-OFF-2532-COMP-0-2FT	3421	0.7710	-7471A	PPB	03 Feb 2017 19:55:48	U	SMPL
I1665-01DUP - 1	ROLL-OFF-2532-COMP-0-2FTDUP	3871	0.8718	-7471A	PPB	03 Feb 2017 19:58:04	U	SMPL
I1665-01MS - 1	ROLL-OFF-2532-COMP-0-2FTMS	20262	4.5413	-7471A	PPB	03 Feb 2017 20:00:20	U	SMPL
I1665-01MSD - 1	ROLL-OFF-2532-COMP-0-2FTMSD	20493	4.5931	-7471A	PPB	03 Feb 2017 20:02:37	U	SMPL
I1665-01LX5 - 1		672	0.1556	-7471A	PPB	03 Feb 2017 20:04:54	U	SMPL
I1665-01A - 1		20654	4.6291	-7471A	PPB	03 Feb 2017 20:07:12	U	SMPL
CCV69 - 1	CCV69	21557	4.8313	-7471A	PPB	03 Feb 2017 20:09:29	U	SMPL
CCB69 - 1	CCB69	-9	0.0031	-7471A	PPB	03 Feb 2017 20:11:47	U	SMPL