

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB85691

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP50397
calibration std. phosphate 1 ppm	WP52706
calibration std. phosphate 0.5 ppm	WP52705
calibration std. phosphate 0.3 ppm	WP52704
calibration std. 0 ppm	WP52701
phosphate ICV std.	WP52708
phosphate CCV std.	WP52707
calibration std. phosphate 0.1 ppm	WP52703
calibration std. phosphate 0.05 ppm	WP52702
Combined reagent	WP52747
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP52746

Intercept: -0.0003Slope: 0.6538Regression: 0.999932

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	02/06/2017	11:30
2	CAL2	0.05	1	50	50	0.031	02/06/2017	11:30
3	CAL3	0.10	1	50	50	0.063	02/06/2017	11:31
4	CAL4	0.30	1	50	50	0.196	02/06/2017	11:31
5	CAL5	0.50	1	50	50	0.332	02/06/2017	11:32
6	CAL6	1.00	1	50	50	0.651	02/06/2017	11:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB85691

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.499	02/06/2017	11:33
2	ICB		1	50	50	0.000	0.000	02/06/2017	11:33
3	CCV1	0.50	1	50	50	0.331	0.507	02/06/2017	11:34
4	CCB1		1	50	50	0.000	0.000	02/06/2017	11:34
5	PB96737BL		1	50	50	0.000	0.000	02/06/2017	11:35
6	PB96737BS	0.50	1	50	50	0.329	0.504	02/06/2017	11:35
7	I1675-01		1	50	50	0.698	1.068	02/06/2017	11:36
8	I1675-01DUP		1	50	50	0.696	1.065	02/06/2017	11:36
9	I1675-01MS	0.50	1	50	50	0.721	1.103	02/06/2017	11:37
10	I1675-01MSD	0.50	1	50	50	0.723	1.106	02/06/2017	11:37
11	I1675-01		5	50	50	0.112	0.172	02/06/2017	11:38
12	I1675-01DUP		5	50	50	0.113	0.173	02/06/2017	11:38
13	I1675-01MS	0.50	5	50	50	0.171	0.262	02/06/2017	11:39
14	I1675-01MSD	0.50	5	50	50	0.172	0.264	02/06/2017	11:39
15	CCV2	0.50	1	50	50	0.330	0.505	02/06/2017	11:40
16	CCB2		1	50	50	0.000	0.000	02/06/2017	11:40