



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB85692

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP52743
COD calibration std. 100 ppm	WP52740
COD calibration std. 10 ppm	WP52739
COD calibration std. 150 ppm	WP52741
COD calibration std. 50 ppm	WP52742
COD calibration std. 0 ppm	WP52738
COD CCV std, 50ppm	WP52744

Temp In(C): <u>150</u>	Date In: <u>02/06/2017</u>	Time In: <u>09:30</u>
Temp Out(C): <u>150</u>	Date Out: <u>02/06/2017</u>	Time Out: <u>11:31</u>

Intercept: 0

Slope: 1

Regression: 1

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	02/06/2017	12:40
2	CAL2	10	1	Water	10.000	02/06/2017	12:40
3	CAL3	50	1	Water	50.000	02/06/2017	12:41
4	CAL4	100	1	Water	100.000	02/06/2017	12:41
5	CAL5	150	1	Water	150.000	02/06/2017	12:42



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB85692

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.000	02/06/2017	12:42
2	ICB				1	Water	0.000	0.000	02/06/2017	12:43
3	CCV1	50			1	Water	49.000	49.000	02/06/2017	12:43
4	CCB1				1	Water	0.000	0.000	02/06/2017	12:44
5	LB85692BL				1	Water	0.000	0.000	02/06/2017	12:44
6	LB85692BS	50			1	Water	48.000	48.000	02/06/2017	12:45
7	I1648-01				1	Water	0.000	0.000	02/06/2017	12:45
8	I1648-01DUP				1	Water	0.000	0.000	02/06/2017	12:46
9	I1648-01MS	50			1	Water	50.000	50.000	02/06/2017	12:47
10	I1648-01MSD	50			1	Water	50.000	50.000	02/06/2017	12:47
11	I1648-02				1	Water	0.000	0.000	02/06/2017	12:48
12	CCV2	50			1	Water	49.000	49.000	02/06/2017	12:48
13	CCB2				1	Water	0.000	0.000	02/06/2017	12:49

WORKLIST(Hardcopy Internal Chain)

2B85692

Worklist Name : COD020316

Worklist ID : 95327

Date : 2/3/2017 2:09:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1648-01	COD	Conc H2SO4 1c	SHAW22	G11	OUTFALL001-170124	01/24/2017	SM5220 D
	Water	I1648-02	COD	Conc H2SO4 1c	SHAW22	G11	OUTFALL001-170131	01/31/2017	SM5220 D

Date/Time 02/06/17 9:25

Received by: IS

Relinquished by: IS

Date/Time 02/06/17 10:20

Received by: IS

Relinquished by: IS