


7470A

LB85715
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37608



A= 0.0000e+000

B= 2.6596e-004

C= 6.0000e-002

Rho= 0.9999287

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	-58	0.000	-58				
0.2	0.200	0.188	-0.012	480	0.0 %	480				
2.5	2.500	2.510	0.010	9213	0.0 %	9213				
5.0	5.000	4.949	-0.051	18381	0.0 %	18381				
7.5	7.500	7.446	-0.054	27772	0.0 %	27772				
10.0	10.000	10.062	0.062	37608	0.0 %	37608				

LB85715 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Feb 2017 09:51:04				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	-58	-	0.0000	7470A	PPB	07 Feb 2017 09:55:46	S
0.2 - 1	50.2	480	-	0.2000	7470A	PPB	07 Feb 2017 09:57:54	S
2.5 - 1	52.5	9213	-	2.5000	7470A	PPB	07 Feb 2017 10:00:03	S
5.0 - 1	55	18381	-	5.0000	7470A	PPB	07 Feb 2017 10:02:11	S
7.5 - 1	57.5	27772	-	7.5000	7470A	PPB	07 Feb 2017 10:04:19	S
10.0 - 1	510	37608	-	10.0000	7470A	PPB	07 Feb 2017 10:06:26	S
ICV20 - 1	ICV20	15207	4.1045	-	7470A	PPB	07 Feb 2017 10:09:11	U
ICB20 - 1	ICB20	-36	0.0504	-	7470A	PPB	07 Feb 2017 10:11:17	U
CCV70 - 1	CCV70	18443	4.9652	-	7470A	PPB	07 Feb 2017 10:13:24	U
CCB70 - 1	CCB70	-31	0.0518	-	7470A	PPB	07 Feb 2017 10:15:30	U
CRA - 1	CRA	537	0.2028	-	7470A	PPB	07 Feb 2017 10:17:37	U
HighStd - 1	HighStd	37570	10.0523	-	7470A	PPB	07 Feb 2017 10:19:43	U
ChkStd - 1	ChkStd	26330	7.0628	-	7470A	PPB	07 Feb 2017 10:21:50	U
PB96751BL - 1	PBW	-54	0.0456	-	7470A	PPB	07 Feb 2017 10:23:57	U
PB96751BS - 1	LCSW	15132	4.0846	-	7470A	PPB	07 Feb 2017 10:26:05	U
PB96751TB - 1	PB96679TB	-12	0.0568	-	7470A	PPB	07 Feb 2017 10:28:13	U
I1660-01 - 1	SB-34-COMP	-14	0.0563	-	7470A	PPB	07 Feb 2017 10:30:21	U
I1660-02 - 1	SB-28-COMP	-20	0.0547	-	7470A	PPB	07 Feb 2017 10:32:29	U
I1660-03 - 1	SB-22-COMP	-8	0.0579	-	7470A	PPB	07 Feb 2017 10:34:38	U
I1660-04 - 1	SB-43-COMP	-8	0.0579	-	7470A	PPB	07 Feb 2017 10:36:45	U
CCV71 - 1	CCV71	18552	4.9942	-	7470A	PPB	07 Feb 2017 10:38:51	U
CCB71 - 1	CCB71	-40	0.0494	-	7470A	PPB	07 Feb 2017 10:40:58	U
I1660-05 - 1	SB-44-COMP	-9	0.0576	-	7470A	PPB	07 Feb 2017 10:43:05	U
I1660-06 - 1	SB-35-COMP	-7	0.0581	-	7470A	PPB	07 Feb 2017 10:45:12	U
I1660-07 - 1	SB-36-COMP	2	0.0605	-	7470A	PPB	07 Feb 2017 10:47:19	U
I1660-08 - 1	SB-37-COMP	-5	0.0587	-	7470A	PPB	07 Feb 2017 10:49:26	U
I1660-09 - 1	SB-38-COMP	-21	0.0544	-	7470A	PPB	07 Feb 2017 10:51:34	U
I1660-10 - 1	SB-39-COMP	-9	0.0576	-	7470A	PPB	07 Feb 2017 10:53:41	U
I1660-11 - 1	SB-41-COMP	-9	0.0576	-	7470A	PPB	07 Feb 2017 10:55:50	U
I1660-12 - 1	SB-42-COMP	1	0.0603	-	7470A	PPB	07 Feb 2017 10:57:58	U
I1687-03 - 1	WC-B6-B7-001	28	0.0674	-	7470A	PPB	07 Feb 2017 11:00:06	U
I1687-03DUP - 1	WC-B6-B7-001DUP	42	0.0712	-	7470A	PPB	07 Feb 2017 11:02:13	U
CCV72 - 1	CCV72	18643	5.0184	-	7470A	PPB	07 Feb 2017 11:04:20	U
CCB72 - 1	CCB72	-30	0.0520	-	7470A	PPB	07 Feb 2017 11:06:27	U
I1687-03MS - 1	WC-B6-B7-001MS	14230	3.8447	-	7470A	PPB	07 Feb 2017 11:08:34	U
I1687-03MSD - 1	WC-B6-B7-001MSD	14176	3.8303	-	7470A	PPB	07 Feb 2017 11:10:41	U
I1687-04 - 1	WC-B25-B30-001	96	0.0855	-	7470A	PPB	07 Feb 2017 11:12:47	U
PB96753BL - 1	PBW	9	0.0624	-	7470A	PPB	07 Feb 2017 11:14:54	U
PB96753BS - 1	LCSW	13876	3.7505	-	7470A	PPB	07 Feb 2017 11:17:02	U
PB96753TB - 1	PB967137B	-5	0.0587	-	7470A	PPB	07 Feb 2017 11:19:09	U
I1690-04 - 1	MH-6-COMP	10	0.0627	-	7470A	PPB	07 Feb 2017 11:21:17	U
I1695-10 - 1	B2-9	276	0.1334	-	7470A	PPB	07 Feb 2017 11:23:25	U
I1695-11 - 1	B2-10	41	0.0709	-	7470A	PPB	07 Feb 2017 11:25:34	U
I1695-12 - 1	B2-11	25	0.0666	-	7470A	PPB	07 Feb 2017 11:27:41	U
CCV73 - 1	CCV73	18627	5.0141	-	7470A	PPB	07 Feb 2017 11:29:48	U
CCB73 - 1	CCB73	-42	0.0488	-	7470A	PPB	07 Feb 2017 11:31:55	U
1697-06 - 1	WC-B10-B17-001	15	0.0640	-	7470A	PPB	07 Feb 2017 11:34:02	U
1697-07 - 1	WC-B26-B27-001	92	0.0845	-	7470A	PPB	07 Feb 2017 11:36:09	U
1697-08 - 1	WC-B18-B19-001	49	0.0730	-	7470A	PPB	07 Feb 2017 11:38:16	U
1697-09 - 1	WC-B23-B24-001	177	0.1071	-	7470A	PPB	07 Feb 2017 11:40:24	U
1697-10 - 1	WC-B20-B22-001	69	0.0784	-	7470A	PPB	07 Feb 2017 11:42:31	U
1698-05 - 1	WC-B8-B12-001	85	0.0826	-	7470A	PPB	07 Feb 2017 11:44:39	U
1698-06 - 1	WC-B4-B9-001	81	0.0815	-	7470A	PPB	07 Feb 2017 11:46:47	U
1698-07 - 1	WC-B13-B14-001	19	0.0651	-	7470A	PPB	07 Feb 2017 11:48:55	U
1698-08 - 1	WC-B11-B16-001	310	0.1424	-	7470A	PPB	07 Feb 2017 11:51:04	U
1698-08DUP - 1	WC-B11-B16-001DUP	327	0.1470	-	7470A	PPB	07 Feb 2017 11:53:11	U
CV74 - 1	CCV74	18402	4.9543	-	7470A	PPB	07 Feb 2017 11:55:18	U
CB74 - 1	CCB74	-23	0.0539	-	7470A	PPB	07 Feb 2017 11:57:25	U
1698-08MS - 1	WC-B11-B16-001MS	14035	3.7928	-	7470A	PPB	07 Feb 2017 11:59:32	U
1698-08MSD - 1	WC-B11-B16-001MSD	13915	3.7609	-	7470A	PPB	07 Feb 2017 12:01:40	U
1687-03LX5 - 1		-34	0.0510	-	7470A	PPB	07 Feb 2017 12:03:46	U
I687-03A - 1		15056	4.0644	-	7470A	PPB	07 Feb 2017 12:05:54	U
I698-08LX5 - 1		31	0.0682	-	7470A	PPB	07 Feb 2017 12:08:01	U
I698-08A - 1		14932	4.0314	-	7470A	PPB	07 Feb 2017 12:10:09	U
CV75 - 1	CCV75	18643	5.0184	-	7470A	PPB	07 Feb 2017 12:12:17	U
CB75 - 1	CCB75	-38	0.0499	-	7470A	PPB	07 Feb 2017 12:14:25	U