

245.1


L1385717

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35452



A= 0.0000e+000

B= 2.8159e-004

C= -2.6682e-002

Rho= 0.9999151

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	1	0.000	1				
0.2	0.200	0.163	-0.037	674	0.0 %	674				
2.5	2.500	2.527	0.027	9069	0.0 %	9069				
5.0	5.000	5.095	0.095	18187	0.0 %	18187				
7.5	7.500	7.485	-0.015	26677	0.0 %	26677				
10.0	10.000	9.956	-0.044	35452	0.0 %	35452				

LB85717 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Feb 2017 13:06:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	1	-	0.0000	245.1	PPB	07 Feb 2017 14:09:06	S	Std
0.2 - 1	SD	674	-	0.2000	245.1	PPB	07 Feb 2017 14:18:33	S	Std
2.5 - 1	SD	9069	-	2.5000	245.1	PPB	07 Feb 2017 14:20:41	S	Std
5.0 - 1	SD	18187	-	5.0000	245.1	PPB	07 Feb 2017 14:22:49	S	Std
7.5 - 1	SD	26677	-	7.5000	245.1	PPB	07 Feb 2017 14:24:57	S	Std
10.0 - 1	SD	35452	-	10.0000	245.1	PPB	07 Feb 2017 14:27:03	S	Std
ICV21 - 1	ICV21	13957	3.9035	-245.1	PPB	PPB	07 Feb 2017 14:29:45	U	SMPL
ICB21 - 1	ICB21	-14	-0.0306	-245.1	PPB	PPB	07 Feb 2017 14:31:51	U	SMPL
CCV76 - 1	CCV76	18132	5.0791	-245.1	PPB	PPB	07 Feb 2017 14:33:57	U	SMPL
CCB76 - 1	CCB76	-2	-0.0272	-245.1	PPB	PPB	07 Feb 2017 14:36:04	U	SMPL
CRA - 1	CRA	724	0.1772	-245.1	PPB	PPB	07 Feb 2017 14:38:10	U	SMPL
HighStd - 1	HighStd	35245	9.8980	-245.1	PPB	PPB	07 Feb 2017 14:40:17	U	SMPL
ChkStd - 1	ChkStd	24294	6.8143	-245.1	PPB	PPB	07 Feb 2017 14:42:24	U	SMPL
PB96800BL - 1	PBW	-17	-0.0315	-245.1	PPB	PPB	07 Feb 2017 14:44:31	U	SMPL
PB96800BS - 1	LCSW	13632	3.8120	-245.1	PPB	PPB	07 Feb 2017 14:46:38	U	SMPL
I1641-01 - 1	013017-DSDC-F-D-M	57	-0.0106	-245.1	PPB	PPB	07 Feb 2017 14:48:46	U	SMPL
I1641-02 - 1	013017-DSDC-F-U-M	69	-0.0073	-245.1	PPB	PPB	07 Feb 2017 14:50:54	U	SMPL
CCV77 - 1	CCV77	17941	5.0253	-245.1	PPB	PPB	07 Feb 2017 14:53:02	U	SMPL
CCB77 - 1	CCB77	-9	-0.0292	-245.1	PPB	PPB	07 Feb 2017 14:55:10	U	SMPL