



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85719

Temp.: 20.1 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/07/2017	13:25
2	CAL2	1	Water			7.01	02/07/2017	13:26
3	CAL3	1	Water			10.03	02/07/2017	13:27
4	ICV	1	Water			7.01	02/07/2017	13:28
5	CCV1	1	Water			2.02	02/07/2017	13:29
6	I1748-01	1	Water			4.23	02/07/2017	13:30
7	I1748-01DUP	1	Water			4.24	02/07/2017	13:31
8	CCV2	1	Water			12.02	02/07/2017	13:32

1385719

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1748

WorkList ID : 95425

Date : 2/7/2017 1:03:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1748-01	pH	Cool 4 deg C	STER14	G51	POOLWATER	02/02/2017	SM 4500-PH B

Date/Time 02/07/17 11:03 AM
Received by: JR
Relinquished by:

Date/Time 02/07/17 2:30 PM
Received by: JR
Relinquished by: JR