



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85721

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/07/2017	13:25
2	CAL2	1	Water			7.01	02/07/2017	13:26
3	CAL3	1	Water			10.03	02/07/2017	13:27
4	ICV	1	Water			7.01	02/07/2017	13:28
5	CCV1	1	Water			2.02	02/07/2017	13:45
6	I1728-04	1	Solid	20.02	20	6.54	02/07/2017	13:46
7	I1731-04	1	Solid	20.01	20	8.32	02/07/2017	13:47
8	I1751-04	1	Solid	20.04	20	7.51	02/07/2017	13:48
9	I1751-04DUP	1	Solid	20.02	20	7.52	02/07/2017	13:50
10	CCV2	1	Water			12.01	02/07/2017	13:51

WORKLIST(Hardcopy Internal Chain)

2B85721

WorkList Name : CORROSIONITY I1751

WorkList ID : 95426

Date : 2/7/2017 1:29:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/09/2017	Solid	I1728-04	Corrosivity	Cool 4 deg C	PSEG01	E42	ROLL-OFF-R2-COMP	02/06/2017	9045C
02/09/2017	Solid	I1731-04	Corrosivity	Cool 4 deg C	PSEG01	E42	ROLL-OFF-1615-COMP	02/06/2017	9045C
02/10/2017	Solid	I1751-04	Corrosivity	Cool 4 deg C	PSEG01	E32	ROLL-OFF-COMP	02/07/2017	9045C

Date/Time 02/07/17 11:29 AM
 Received by: JR
 Relinquished by: JR

Date/Time 02/07/17 5:00 PM
 Received by: JR
 Relinquished by: JR