



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85722

Temp.: 24.3 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2083
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/07/2017	14:00
2	CAL2	1	Water			7.01	02/07/2017	14:01
3	CAL3	1	Water			10.02	02/07/2017	14:02
4	ICV	1	Water			7.02	02/07/2017	14:03
5	CCV1	1	Water			2.01	02/07/2017	14:04
6	I1745-03	1	Solid	20.02	20	6.88	02/07/2017	14:05
7	I1745-03DUP	1	Solid	20.04	20	6.89	02/07/2017	14:06
8	CCV2	1	Water			12.01	02/07/2017	14:07

1885722

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I1745

WorkList ID : 95429

Date : 2/7/2017 1:34:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/03/2017	Solid	I1745-03	pH	Cool 4 deg C	CYRU01	G31	SB-1	02/03/2017	9045C

Date/Time 02/07/17 1:34 PM
Received by: JF
Relinquished by: JF

Date/Time 02/07/17 3:25 PM
Received by: JF
Relinquished by: JF