

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/8/17
 Analyst : AP
 Data File : lb85724.csv

Approved By : AK
 Approved Date : 2-7-17
 Worksheet # : _____

M 3000 - Inorganic Anions - 16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Bromide											
STD1	STD1				W	1/16/17					
Bromide		PASS	0.000		0.000						mg/L
STD2	STD2				W	1/16/17					
Bromide		PASS	0.913		0.913						mg/L
STD3	STD3				W	1/16/17					
Bromide		PASS	2.270		2.27						mg/L
STD4	STD4				W	1/16/17					
Bromide		PASS	4.082		4.08						mg/L
STD5	STD5				W	1/16/17					
Bromide		PASS	9.756		9.76						mg/L
STD6	STD6				W	1/16/17					
Bromide		PASS	19.527		19.5						mg/L
STD7	STD7				W	1/16/17					
Bromide		PASS	24.510		24.5						mg/L
STD8	STD8				W	1/16/17					
Bromide		PASS	50.442		50.4						mg/L
ICV1	ICV1				W	1/16/17					
Bromide		PASS	25.395		25.40	102.0	90	110			mg/L
ICB1	ICB1				W	1/16/17					
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L
CCV1	CCV1				W	2/7/17					
Bromide		PASS	24.276		24.30	97.0	90	110			mg/L
CCB1	CCB1				W	2/7/17					
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L
LB85724BLW	LB85724BLW				W	2/7/17					
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L
LB85724BSW	LB85724BSW				W	2/7/17					
Bromide		PASS	24.228		24.20	97.0	90	110			mg/L
I1748-01	POOLWATER		1		W	2/7/17					
Bromide		PASS	0.000		0.000						mg/L
I1748-01D	POOLWATERD		1		W	2/7/17					
Bromide		PASS	0.000		0.00				0	20	mg/L
I1748-01S	POOLWATERS		1		W	2/7/17					
Bromide		PASS	24.087		24.10	96.0	80	120			mg/L
CCV2	CCV2				W	2/7/17					
Bromide		PASS	24.340		24.30	97.0	90	110			mg/L
CCB2	CCB2				W	2/7/17					
Bromide		PASS	0.000		0.000			+/-0.0660			mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/8/17
 Analyst : AP
 Data File : lb85724.csv

Approved By : AK
 Approved Date : 2-10-17
 Worksheet # : _____

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
Parameter			PPB		Final Conc	%Rec	Method	Method			Line 1
							LCL	UCL	RPD	Max RPD	Units
											Line 2
Chloride											
STD1	STD1				W	1/16/17					
Chloride		PASS	0.000		0.000						mg/L
STD2	STD2				W	1/16/17					
Chloride		PASS	0.357		0.357						mg/L
STD3	STD3				W	1/16/17					
Chloride		PASS	0.765		0.765						mg/L
STD4	STD4				W	1/16/17					
Chloride		PASS	1.259		1.26						mg/L
STD5	STD5				W	1/16/17					
Chloride		PASS	2.867		2.87						mg/L
STD6	STD6				W	1/16/17					
Chloride		PASS	5.722		5.72						mg/L
STD7	STD7				W	1/16/17					
Chloride		PASS	7.248		7.25						mg/L
STD8	STD8				W	1/16/17					
Chloride		PASS	15.231		15.2						mg/L
ICV1	ICV1				W	1/16/17					
Chloride		PASS	7.295		7.30	97.0	90	110			mg/L
ICB1	ICB1				W	1/16/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
CCV1	CCV1				W	2/7/17					
Chloride		PASS	7.074		7.07	94.0	90	110			mg/L
CCB1	CCB1				W	2/7/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
LB85724BLW	LB85724BLW				W	2/7/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
LB85724BSW	LB85724BSW				W	2/7/17					
Chloride		PASS	7.156		7.16	95.0	90	110			mg/L
I1748-01	POOLWATER		I		W	2/7/17					
Chloride		PASS	0.322		0.322						mg/L
I1748-01D	POOLWATERD		I		W	2/7/17					
Chloride		PASS	0.366		0.37				13.9	20	mg/L
I1748-01S	POOLWATERS		I		W	2/7/17					
Chloride		PASS	7.302		7.30	93.0	80	120			mg/L
CCV2	CCV2				W	2/7/17					
Chloride		PASS	7.206		7.21	96.0	90	110			mg/L
CCB2	CCB2				W	2/7/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/8/17
 Analyst : AP
 Data File : lb85724.csv

Approved By : AK
 Approved Date : 2-10-17
 Worksheet # : _____

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Fluoride											
STD1	STD1			W	1/16/17						
Fluoride		PASS		0.000	0.000					mg/L	
STD2	STD2			W	1/16/17						
Fluoride		PASS		0.255	0.255					mg/L	
STD3	STD3			W	1/16/17						
Fluoride		PASS		0.470	0.470					mg/L	
STD4	STD4			W	1/16/17						
Fluoride		PASS		0.845	0.845					mg/L	
STD5	STD5			W	1/16/17						
Fluoride		PASS		1.854	1.85					mg/L	
STD6	STD6			W	1/16/17						
Fluoride		PASS		3.888	3.89					mg/L	
STD7	STD7			W	1/16/17						
Fluoride		PASS		4.862	4.86					mg/L	
STD8	STD8			W	1/16/17						
Fluoride		PASS		10.125	10.1					mg/L	
ICV1	ICV1			W	1/16/17						
Fluoride		PASS		4.693	4.69	94.0	110			mg/L	
ICB1	ICB1			W	1/16/17						
Fluoride		PASS		0.000	0.000		+/-0.0430			mg/L	
CCV1	CCV1			W	2/7/17						
Fluoride		PASS		4.978	4.98	100.0	90	110		mg/L	
CCB1	CCB1			W	2/7/17						
Fluoride		PASS		0.000	0.000		+/-0.0430			mg/L	
LB85724BLW	LB85724BLW			W	2/7/17						
Fluoride		PASS		0.000	0.000		+/-0.0430			mg/L	
LB85724BSW	LB85724BSW			W	2/7/17						
Fluoride		PASS		4.884	4.88	98.0	90	110		mg/L	
I1748-01	POOLWATER		1	W	2/7/17						
Fluoride		PASS		0.000	0.000					mg/L	
I1748-01D	POOLWATERD		1	W	2/7/17						
Fluoride		PASS		0.000	0.00			0	20	mg/L	
I1748-01S	POOLWATERS		1	W	2/7/17						
Fluoride		PASS		4.858	4.86	97.0	80	120		mg/L	
CCV2	CCV2			W	2/7/17						
Fluoride		PASS		5.015	5.01	100.0	90	110		mg/L	
CCB2	CCB2			W	2/7/17						
Fluoride		PASS		0.000	0.000		+/-0.0430			mg/L	