

7471A


LB 85726

INSTRUMENT ID: CV4

Linear

 μ Abs.:

44702



A= 0.0000e+000

B= 2.2326e-004

C= -5.2736e-002

Rho= 0.9998695

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.041	-0.041	52	0.000	52				
0.2	0.200	0.224	0.024	1240	0.0 %	1240				
2.5	2.500	2.511	0.011	11484	0.0 %	11484				
5.0	5.000	4.967	-0.033	22482	0.0 %	22482				
7.5	7.500	7.612	0.112	34331	0.0 %	34331				
10.0	10.000	9.927	-0.073	44702	0.0 %	44702				

LB85726 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 07 Feb 2017 15:39:33

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.0000	7471A	PPB	07 Feb 2017 15:40:49	S	Std
0.2 - 1	50.2	1240	-	0.2000	7471A	PPB	07 Feb 2017 15:43:06	S	Std
2.5 - 1	52.5	11484	-	2.5000	7471A	PPB	07 Feb 2017 15:45:24	S	Std
5.0 - 1	55	22482	-	5.0000	7471A	PPB	07 Feb 2017 15:47:41	S	Std
7.5 - 1	57.5	34331	-	7.5000	7471A	PPB	07 Feb 2017 15:49:59	S	Std
10.0 - 1	57.5	44702	-	10.0000	7471A	PPB	07 Feb 2017 15:54:50	S	Std
ICV22 - 1	ICV22	17280	3.8052	-	7471A	PPB	07 Feb 2017 15:58:44	U	SMPL
ICB22 - 1	ICB22	11	-0.0503	-	7471A	PPB	07 Feb 2017 16:00:59	U	SMPL
CCV78 - 1	CCV78	23477	5.1887	-	7471A	PPB	07 Feb 2017 16:03:15	U	SMPL
CCB78 - 1	CCB78	-5	-0.0539	-	7471A	PPB	07 Feb 2017 16:05:30	U	SMPL
CRA - 1	CRA	1188	0.2125	-	7471A	PPB	07 Feb 2017 16:07:46	U	SMPL
HighStd - 1	HighStd	45249	10.0495	-	7471A	PPB	07 Feb 2017 16:10:02	U	SMPL
ChkStd - 1	ChkStd	32180	7.1317	-	7471A	PPB	07 Feb 2017 16:12:18	U	SMPL
PB96768BL - 1	PBS	43	-0.0431	-	7471A	PPB	07 Feb 2017 16:14:34	U	SMPL
PB96768BS - 1	LCSS	15557	3.4205	-	7471A	PPB	07 Feb 2017 16:16:51	U	SMPL
I1685-01 - 1	N-1A	16100	3.5417	-	7471A	PPB	07 Feb 2017 16:19:07	U	SMPL
I1690-01 - 1	MH-6-COMP	33169	7.3525	-	7471A	PPB	07 Feb 2017 16:21:25	U	SMPL
I1692-01 - 1	ROLL-OFF-15-40-COMP	2297	0.4601	-	7471A	PPB	07 Feb 2017 16:23:42	U	SMPL
I1692-04 - 1	ROLL-OFF-20-41-COMP	1428	0.2661	-	7471A	PPB	07 Feb 2017 16:26:00	U	SMPL
I1695-03 - 1	B2-9	136522	30.4269	-	7471A	PPB	07 Feb 2017 16:28:16	U	SMPL
CCV79 - 1	CCV79	23441	5.1807	-	7471A	PPB	07 Feb 2017 16:30:31	U	SMPL
CCB79 - 1	CCB79	-12	-0.0554	-	7471A	PPB	07 Feb 2017 16:32:48	U	SMPL
I1695-06 - 1	B2-10	121192	27.0044	-	7471A	PPB	07 Feb 2017 16:35:05	U	SMPL
I1695-09 - 1	B2-11	139503	31.0925	-	7471A	PPB	07 Feb 2017 16:37:20	U	SMPL
I1696-01 - 1	FK-BPM-01-A5B-A6B-A7B-A8B	289	0.0118	-	7471A	PPB	07 Feb 2017 16:42:28	U	SMPL
I1696-01DUP - 1	FK-BPM-01-A5B-A6B-A7B-A8BDUP	297	0.0136	-	7471A	PPB	07 Feb 2017 16:44:44	U	SMPL
I1696-01MS - 1	FK-BPM-01-A5B-A6B-A7B-A8BMS	16608	3.6551	-	7471A	PPB	07 Feb 2017 16:47:01	U	SMPL
I1696-01MSD - 1	FK-BPM-01-A5B-A6B-A7B-A8BMSD	16401	3.6089	-	7471A	PPB	07 Feb 2017 16:49:18	U	SMPL
I1699-01 - 1	F-1	1632	0.3116	-	7471A	PPB	07 Feb 2017 16:51:35	U	SMPL
I1699-03 - 1	F-2	1238	0.2237	-	7471A	PPB	07 Feb 2017 16:53:52	U	SMPL
I1701-01 - 1	C-1	557	0.0716	-	7471A	PPB	07 Feb 2017 16:56:10	U	SMPL
I1704-01 - 1	FK-SF-01-002-A	1605	0.3056	-	7471A	PPB	07 Feb 2017 16:58:26	U	SMPL
CCV80 - 1	CCV80	23749	5.2494	-	7471A	PPB	07 Feb 2017 17:00:43	U	SMPL
CCB80 - 1	CCB80	1	-0.0525	-	7471A	PPB	07 Feb 2017 17:02:59	U	SMPL
I1704-03 - 1	FK-SF-01-002-B	1755	0.3391	-	7471A	PPB	07 Feb 2017 17:05:15	U	SMPL
I1706-02 - 1	WC-SCW-COMP-02	397	0.0359	-	7471A	PPB	07 Feb 2017 17:07:32	U	SMPL
I1706-04 - 1	WC-SCW-COMP-04	393	0.0350	-	7471A	PPB	07 Feb 2017 17:09:48	U	SMPL
I1712-01 - 1	DRUMS-SOIL-ROW-COMP	810	0.1281	-	7471A	PPB	07 Feb 2017 17:12:05	U	SMPL
I1728-01 - 1	ROLL-OFF-R2-COMP	1060	0.1839	-	7471A	PPB	07 Feb 2017 17:14:21	U	SMPL
I1731-01 - 1	ROLL-OFF-1615-COMP	1341	0.2467	-	7471A	PPB	07 Feb 2017 17:16:37	U	SMPL
I1696-01LX5 - 1		87	-0.0333	-	7471A	PPB	07 Feb 2017 17:18:55	U	SMPL
I1696-01A - 1		18959	4.1800	-	7471A	PPB	07 Feb 2017 17:21:12	U	SMPL
CCV81 - 1	CCV81	22739	5.0239	-	7471A	PPB	07 Feb 2017 17:30:17	U	SMPL
CCB81 - 1	CCB81	63	-0.0387	-	7471A	PPB	07 Feb 2017 17:32:33	U	SMPL
I1695-03DLX5 - 1		28065	6.2130	-	7471A	PPB	07 Feb 2017 17:34:49	U	SMPL
I1695-06DLX5 - 1		24786	5.4809	-	7471A	PPB	07 Feb 2017 17:37:06	U	SMPL
I1695-09DLX5 - 1		28674	6.3490	-	7471A	PPB	07 Feb 2017 17:39:22	U	SMPL
CCV82 - 1	CCV82	22257	4.9163	-	7471A	PPB	07 Feb 2017 17:41:39	U	SMPL
CCB82 - 1	CCB82	-6	-0.0541	-	7471A	PPB	07 Feb 2017 17:43:55	U	SMPL