

Run Number: LB85740

DATE & TIME IN: 02/08/2017 10:30

OVEN TEMP: 104 °C

DATE & TIME OUT: 02/08/2017 14:40

OVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB85740BL	LB85740BL	1.5206	1.5206	100	1.5206	1.5206	1.5206	0	0
2	LB85740BS	LB85740BS	1.3641	1.3641	100	1.4187	1.4187	1.4187	0.0546	546
3	I1739-01	020217-PR-E-D-M	1.4058	1.4058	400	1.4171	1.4171	1.4171	0.0113	28.3
4	I1739-02	020217-PR-E-U-M	1.4051	1.4051	400	1.4184	1.4184	1.4184	0.0133	33.2
5	I1740-01	020317-PR-E-D-M	1.4096	1.4096	400	1.4234	1.4234	1.4234	0.0138	34.5
6	I1740-02	020317-PR-E-U-M	1.413	1.413	400	1.4261	1.4261	1.4261	0.0131	32.8
7	I1741-01	020617-PR-E-D-M	1.4069	1.4069	400	1.4205	1.4205	1.4205	0.0136	34
8	I1741-02	020617-PR-E-U-M	1.4142	1.4142	400	1.4223	1.4223	1.4223	0.0081	20.2
9	I1742-01	DRGFB-WQ-033-AD-A	1.4142	1.4142	400	1.4295	1.4295	1.4295	0.0153	38.3
10	I1742-02	DRGFB-WQ-036-BG-A	1.4098	1.4098	400	1.4263	1.4263	1.4263	0.0165	41.3
11	I1744-02	KY038PS027-170206	1.402	1.402	800	1.4032	1.4032	1.4032	0.0012	1.5
12	I1747-01	001M-WILLETS-PTBLVD (JAN)	1.4042	1.4042	300	1.4182	1.4182	1.4182	0.014	46.7
13	I1747-02	002M-35THAVE (JAN)	1.4149	1.4149	200	1.4256	1.4256	1.4256	0.0107	53.5
14	I1747-02DUP	002M-35THAVE (JAN) DUP	1.4017	1.4017	200	1.4128	1.4128	1.4128	0.0111	55.5
15	I1756-02	OUTFALL001-170207	1.4102	1.4102	800	1.4107	1.4107	1.4107	0.0005	0.6

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

$$\text{Weight (g)} = C - B$$

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I1740

WorkList ID : 95454

Date : 2/8/2017 8:33:30 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1739-01	TSS	Cool 4 deg C	HDR103	G22	020217-PR-E-D-M	02/02/2017	SM2540 D
	Water	I1739-02	TSS	Cool 4 deg C	HDR103	G22	020217-PR-E-U-M	02/02/2017	SM2540 D
	Water	I1740-01	TSS	Cool 4 deg C	HDR103	G22	020317-PR-E-D-M	02/03/2017	SM2540 D
	Water	I1740-02	TSS	Cool 4 deg C	HDR103	G22	020317-PR-E-U-M	02/03/2017	SM2540 D
02/13/2017	Water	I1741-01	TSS	Cool 4 deg C	HDR103	G22	020617-PR-E-D-M	02/06/2017	SM2540 D
02/13/2017	Water	I1741-02	TSS	Cool 4 deg C	HDR103	G22	020617-PR-E-U-M	02/06/2017	SM2540 D
	Water	I1742-01	TSS	Cool 4 deg C	JAYC01	G42	DRGFB-WQ-033-AD-A	02/06/2017	SM2540 D
	Water	I1742-02	TSS	Cool 4 deg C	JAYC01	G42	DRGFB-WQ-036-BG-A	02/07/2017	SM2540 D
	Water	I1744-02	TSS	Cool 4 deg C	SHAW22	G51	KY038PS027-170206	02/06/2017	SM2540 D
	Water	I1747-01	TSS	Cool 4 deg C	TULL01	G51	001M-WILLETS-PTBLVD(JAN	02/02/2017	SM2540 D
	Water	I1747-02	TSS	Cool 4 deg C	TULL01	G51	002M-35THAVE(JAN)	02/02/2017	SM2540 D
	Water	I1756-02	TSS	Cool 4 deg C	SHAW22	H11	OUTFALL001-170207	02/07/2017	SM2540 D

Date/Time 02/08/17 8:33 AM

Received by: SP

Relinquished by: CP

Date/Time 02/08/17 11:35 AM

Received by: CP

Relinquished by: SP