


7470A

LB 85753
(INSTRUMENT ID: CV1)

Linear

 μ Abs.:

37236



A= 0.0000e+000

B= 2.6953e-004

C= 2.9904e-002

Rho= 0.9998612

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-2	0.000	-2				
0.2	0.200	0.213	0.013	679	0.0 %	679				
2.5	2.500	2.449	-0.051	8974	0.0 %	8974				
5.0	5.000	5.052	0.052	18633	0.0 %	18633				
7.5	7.500	7.391	-0.109	27310	0.0 %	27310				
10.0	10.000	10.066	0.066	37236	0.0 %	37236				

LB85753 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Feb 2017 16:57:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	7470A	PPB	08 Feb 2017 17:00:33	S	Std
0.2 - 1	50.2	679	-	0.2000	7470A	PPB	08 Feb 2017 17:02:41	S	Std
2.5 - 1	525	8974	-	2.5000	7470A	PPB	08 Feb 2017 17:04:50	S	Std
5.0 - 1	55	18633	-	5.0000	7470A	PPB	08 Feb 2017 17:06:58	S	Std
7.5 - 1	57.5	27310	-	7.5000	7470A	PPB	08 Feb 2017 17:09:06	S	Std
10.0 - 1	510	37236	-	10.0000	7470A	PPB	08 Feb 2017 17:11:13	S	Std
ICV23 - 1	ICV23	14001	3.8036	-7470A	PPB	08 Feb 2017 17:14:02	U	SMPL	
ICB23 - 1	ICB23	-27	0.0226	-7470A	PPB	08 Feb 2017 17:16:08	U	SMPL	
CCV89 - 1	CCV89	18347	4.9750	-7470A	PPB	08 Feb 2017 17:18:15	U	SMPL	
CCB89 - 1	CCB89	-35	0.0205	-7470A	PPB	08 Feb 2017 17:20:21	U	SMPL	
CRA - 1	CRA	665	0.2091	-7470A	PPB	08 Feb 2017 17:22:28	U	SMPL	
HighStd - 1	HighStd	36420	9.8462	-7470A	PPB	08 Feb 2017 17:24:34	U	SMPL	
ChkStd - 1	ChkStd	26744	7.2382	-7470A	PPB	08 Feb 2017 17:26:41	U	SMPL	
PB96836BL - 1	PBW	-48	0.0170	-7470A	PPB	08 Feb 2017 17:28:48	U	SMPL	
PB96836BS - 1	LCSW	12840	3.4907	-7470A	PPB	08 Feb 2017 17:30:56	U	SMPL	
PB96836TB - 1	PB96749TB	-24	0.0234	-7470A	PPB	08 Feb 2017 17:33:03	U	SMPL	
I1692-02 - 1	ROLL-OFF-15-40-COMP	-12	0.0267	-7470A	PPB	08 Feb 2017 17:35:12	U	SMPL	
I1692-02DUP - 1	ROLL-OFF-15-40-COMPDUP	-14	0.0261	-7470A	PPB	08 Feb 2017 17:37:20	U	SMPL	
I1692-02MS - 1	ROLL-OFF-15-40-COMPMS	13194	3.5861	-7470A	PPB	08 Feb 2017 17:39:29	U	SMPL	
I1692-02MSD - 1	ROLL-OFF-15-40-COMPMSD	13900	3.7764	-7470A	PPB	08 Feb 2017 17:41:36	U	SMPL	
CCV90 - 1	CCV90	18419	4.9944	-7470A	PPB	08 Feb 2017 17:43:43	U	SMPL	
CCB90 - 1	CCB90	-30	0.0218	-7470A	PPB	08 Feb 2017 17:45:49	U	SMPL	
I1692-05 - 1	ROLL-OFF-20-41-COMP	17	0.0345	-7470A	PPB	08 Feb 2017 17:47:56	U	SMPL	
I1706-09 - 1	WC-SCW-COMP-02	-13	0.0264	-7470A	PPB	08 Feb 2017 17:50:03	U	SMPL	
I1706-11 - 1	WC-SCW-COMP-04	-11	0.0269	-7470A	PPB	08 Feb 2017 17:52:10	U	SMPL	
I1709-03 - 1	SS-1	0	0.0299	-7470A	PPB	08 Feb 2017 17:54:17	U	SMPL	
I1712-04 - 1	DRUMS-SOIL-ROW-COMP	36	0.0396	-7470A	PPB	08 Feb 2017 17:56:25	U	SMPL	
I1728-04 - 1	ROLL-OFF-R2-COMP	-10	0.0272	-7470A	PPB	08 Feb 2017 17:58:32	U	SMPL	
I1731-04 - 1	ROLL-OFF-1615-COMP	-9	0.0275	-7470A	PPB	08 Feb 2017 18:00:40	U	SMPL	
PB96839BL - 1	PBW	-16	0.0256	-7470A	PPB	08 Feb 2017 18:02:48	U	SMPL	
PB96839BS - 1	LCSW	13201	3.5880	-7470A	PPB	08 Feb 2017 18:07:57	U	SMPL	
PB96839TB - 1	PB96807TB	-25	0.0232	-7470A	PPB	08 Feb 2017 18:10:03	U	SMPL	
CCV91 - 1	CCV91	18741	5.0812	-7470A	PPB	08 Feb 2017 18:12:10	U	SMPL	
CCB91 - 1	CCB91	-30	0.0218	-7470A	PPB	08 Feb 2017 18:14:17	U	SMPL	
I1751-04 - 1	ROLL-OFF-COMP	2	0.0304	-7470A	PPB	08 Feb 2017 18:16:23	U	SMPL	
I1751-04DUP - 1	ROLL-OFF-COMPDUP	-2	0.0294	-7470A	PPB	08 Feb 2017 18:18:30	U	SMPL	
I1751-04MS - 1	ROLL-OFF-COMPMS	12708	3.4551	-7470A	PPB	08 Feb 2017 18:20:37	U	SMPL	
I1751-04MSD - 1	ROLL-OFF-COMPMSD	13410	3.6443	-7470A	PPB	08 Feb 2017 18:25:18	U	SMPL	
I1692-02LX5 - 1		-28	0.0224	-7470A	PPB	08 Feb 2017 18:27:26	U	SMPL	
I1692-02A - 1		14109	3.8327	-7470A	PPB	08 Feb 2017 18:29:33	U	SMPL	
I1751-04LX5 - 1		-25	0.0232	-7470A	PPB	08 Feb 2017 18:31:42	U	SMPL	
I1751-04A - 1		14151	3.8440	-7470A	PPB	08 Feb 2017 18:33:50	U	SMPL	
CCV92 - 1	CCV92	18671	5.0623	-7470A	PPB	08 Feb 2017 18:35:59	U	SMPL	
CCB92 - 1	CCB92	-28	0.0224	-7470A	PPB	08 Feb 2017 18:38:05	U	SMPL	