


ISM02.3

LB85757
INSTRUMENT ID: CV1

Linear

 μ Abs.:

49183



A= 0.0000e+000

B= 2.0372e-004

C= 1.4955e-002

Rho= 0.9999284

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.028	0.028	62	0.000	62					-
0.2	0.200	0.231	0.031	1060	0.0 %	1060					16
2.5	2.500	2.478	-0.022	12089	0.0 %	12089					-1
5.0	5.000	4.911	-0.089	24032	0.0 %	24032					-2
7.5	7.500	7.518	0.018	36831	0.0 %	36831					0
10.0	10.000	10.035	0.035	49183	0.0 %	49183					0

MB
2/8/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	62	-		0 ISM02.3	1	2/8/2017 18:54	Std
	0.2 S0.2	1060	-		0.2 ISM02.3	1	2/8/2017 18:56	Std
	2.5 S2.5	12089	-		2.5 ISM02.3	1	2/8/2017 18:58	Std
	5 S5	24032	-		5 ISM02.3	1	2/8/2017 19:00	Std
	7.5 S7.5	36831	-		7.5 ISM02.3	1	2/8/2017 19:02	Std
	10 S10	49183	-		10 ISM02.3	1	2/8/2017 19:05	Std
ICV	ICV	18731	3.8309 -		ISM02.3	1	2/8/2017 19:07	SMPL
ICB	ICB	34	0.0219 -		ISM02.3	1	2/8/2017 19:09	SMPL
CCV083	CCV083	24248	4.9548 -		ISM02.3	1	2/8/2017 19:12	SMPL
CCB083	CCB083	35	0.0221 -		ISM02.3	1	2/8/2017 19:14	SMPL
PB96763BL	PBS004	26	0.0203 -		ISM02.3	1	2/8/2017 19:16	SMPL
I1459-01	MBDNQ1	5866	1.21 -		ISM02.3	1	2/8/2017 19:18	SMPL
I1459-02	MBDP83	330	0.0822 -		ISM02.3	1	2/8/2017 19:20	SMPL
I1459-03	MBDP83D	333	0.0828 -		ISM02.3	1	2/8/2017 19:22	SMPL
I1459-04	MBDP83S	12444	2.5501 -		ISM02.3	1	2/8/2017 19:24	SMPL
I1459-05	MBDP84	448	0.1062 -		ISM02.3	1	2/8/2017 19:26	SMPL
I1459-06	MBDP85	2343	0.4923 -		ISM02.3	1	2/8/2017 19:29	SMPL
I1459-07	MBDP86	1371	0.2943 -		ISM02.3	1	2/8/2017 19:31	SMPL
I1459-08	MBDP87	1561	0.333 -		ISM02.3	1	2/8/2017 19:33	SMPL
I1459-09	MBDP88	1515	0.3236 -		ISM02.3	1	2/8/2017 19:35	SMPL
I1459-10	MBDP89	1458	0.312 -		ISM02.3	1	2/8/2017 19:37	SMPL
I1459-11	MBDP90	1066	0.2321 -		ISM02.3	1	2/8/2017 19:39	SMPL
I1459-12	MBDP91	240	0.0638 -		ISM02.3	1	2/8/2017 19:41	SMPL
I1459-13	MBDP92	155	0.0465 -		ISM02.3	1	2/8/2017 19:43	SMPL
I1459-14	MBDP93	380	0.0924 -		ISM02.3	1	2/8/2017 19:45	SMPL
I1459-15	MBDP94	598	0.1368 -		ISM02.3	1	2/8/2017 19:48	SMPL
I1459-16	MBDP95	286	0.0732 -		ISM02.3	1	2/8/2017 19:50	SMPL
I1459-17	MBDP96	306	0.0773 -		ISM02.3	1	2/8/2017 19:52	SMPL
I1459-18	MBDP97	4909	1.015 -		ISM02.3	1	2/8/2017 19:54	SMPL
I1459-19	MBDP98	1199	0.2592 -		ISM02.3	1	2/8/2017 19:56	SMPL
CCV084	CCV084	24112	4.9271 -		ISM02.3	1	2/8/2017 19:58	SMPL
CCB084	CCB084	36	0.0223 -		ISM02.3	1	2/8/2017 20:00	SMPL
I1459-20	MBDP99	946	0.2077 -		ISM02.3	1	2/8/2017 20:03	SMPL
PB96764BL	PBS004	-8	0.0133 -		ISM02.3	1	2/8/2017 20:05	SMPL
I1462-01	MBDPA1	784	0.1747 -		ISM02.3	1	2/8/2017 20:07	SMPL
I1462-02	MBDPA1D	789	0.1757 -		ISM02.3	1	2/8/2017 20:09	SMPL
I1462-03	MBDPA1S	12900	2.643 -		ISM02.3	1	2/8/2017 20:11	SMPL
I1462-04	MBDPA2	262	0.0683 -		ISM02.3	1	2/8/2017 20:13	SMPL
PB96765BL	PBS004	9	0.0168 -		ISM02.3	1	2/8/2017 20:15	SMPL
I1463-01	MBDMX5	992	0.217 -		ISM02.3	1	2/8/2017 20:17	SMPL
I1463-02	MBDMX5D	974	0.2134 -		ISM02.3	1	2/8/2017 20:19	SMPL
I1463-03	MBDMX5S	12926	2.6483 -		ISM02.3	1	2/8/2017 20:22	SMPL
I1463-04	MBDMX6	713	0.1602 -		ISM02.3	1	2/8/2017 20:24	SMPL
I1463-05	MBDMX7	282	0.0724 -		ISM02.3	1	2/8/2017 20:26	SMPL
I1463-06	MBDMX8	748	0.1673 -		ISM02.3	1	2/8/2017 20:28	SMPL
I1463-07	MBDMX9	611	0.1394 -		ISM02.3	1	2/8/2017 20:30	SMPL

INSTRUMENT ID : CV1

I1463-08	MBDMY0	220	0.0598 -	ISM02.3	1	2/8/2017 20:32	SMPL
I1463-09	MBDMY1	418	0.1001 -	ISM02.3	1	2/8/2017 20:34	SMPL
I1463-10	MBDMY2	376	0.0916 -	ISM02.3	1	2/8/2017 20:36	SMPL
I1463-11	MBDMY3	180	0.0516 -	ISM02.3	1	2/8/2017 20:39	SMPL
I1463-12	MBDMY4	634	0.1441 -	ISM02.3	1	2/8/2017 20:41	SMPL
I1463-13	MBDMY5	278	0.0716 -	ISM02.3	1	2/8/2017 20:43	SMPL
CCV085	CCV085	23692	4.8416 -	ISM02.3	1	2/8/2017 20:45	SMPL
CCB085	CCB085	31	0.0213 -	ISM02.3	1	2/8/2017 20:47	SMPL
I1463-14	MBDMY6	269	0.0698 -	ISM02.3	1	2/8/2017 20:49	SMPL
I1463-15	MBDMY7	343	0.0848 -	ISM02.3	1	2/8/2017 20:51	SMPL
I1463-16	MBDMY8	991	0.2168 -	ISM02.3	1	2/8/2017 20:54	SMPL
I1463-17	MBDMY9	182	0.052 -	ISM02.3	1	2/8/2017 20:56	SMPL
I1463-18	MBDMZ0	391	0.0946 -	ISM02.3	1	2/8/2017 20:58	SMPL
PB96766BL	PBS004	124	0.0402 -	ISM02.3	1	2/8/2017 21:00	SMPL
I1501-01	MBDMT3	250	0.0659 -	ISM02.3	1	2/8/2017 21:02	SMPL
I1501-02	MBDMT3D	262	0.0683 -	ISM02.3	1	2/8/2017 21:04	SMPL
I1501-03	MBDMT3S	12110	2.482 -	ISM02.3	1	2/8/2017 21:06	SMPL
I1501-04	MBDMT4	443	0.1052 -	ISM02.3	1	2/8/2017 21:08	SMPL
I1501-05	MBDMT5	258	0.0675 -	ISM02.3	1	2/8/2017 21:10	SMPL
I1501-06	MBDMT6	308	0.0777 -	ISM02.3	1	2/8/2017 21:13	SMPL
I1501-07	MBDMT7	296	0.0753 -	ISM02.3	1	2/8/2017 21:15	SMPL
I1501-08	MBDMT8	245	0.0649 -	ISM02.3	1	2/8/2017 21:17	SMPL
I1501-09	MBDMT9	929	0.2042 -	ISM02.3	1	2/8/2017 21:19	SMPL
I1501-10	MBDMW0	298	0.0757 -	ISM02.3	1	2/8/2017 21:21	SMPL
I1501-11	MBDMW1	210	0.0577 -	ISM02.3	1	2/8/2017 21:23	SMPL
I1501-12	MBDMW2	246	0.0651 -	ISM02.3	1	2/8/2017 21:25	SMPL
I1501-13	MBDMW3	308	0.0777 -	ISM02.3	1	2/8/2017 21:27	SMPL
I1501-14	MBDMW4	883	0.1948 -	ISM02.3	1	2/8/2017 21:30	SMPL
CCV086	CCV086	23572	4.8171 -	ISM02.3	1	2/8/2017 21:32	SMPL
CCB086	CCB086	30	0.0211 -	ISM02.3	1	2/8/2017 21:34	SMPL
I1501-15	MBDMW5	274	0.0708 -	ISM02.3	1	2/8/2017 21:36	SMPL
PB96767BL	PBS004	26	0.0203 -	ISM02.3	1	2/8/2017 21:38	SMPL
I1618-01	ME5BQ8	10471	2.1481 -	ISM02.3	1	2/8/2017 21:40	SMPL
I1618-02	ME5BQ9	3933	0.8162 -	ISM02.3	1	2/8/2017 21:42	SMPL
I1618-03	ME5BR0	1339	0.2877 -	ISM02.3	1	2/8/2017 21:45	SMPL
I1618-04	ME5BR1	6579	1.3552 -	ISM02.3	1	2/8/2017 21:47	SMPL
I1618-05	ME5BR2	5708	1.1778 -	ISM02.3	1	2/8/2017 21:49	SMPL
I1618-06	ME5BR3	89222	18.1915 -	ISM02.3	1	2/8/2017 21:51	SMPL
I1618-07	ME5BR4	64752	13.2064 -	ISM02.3	1	2/8/2017 21:53	SMPL
I1618-08	ME5BR5	13769	2.82 -	ISM02.3	1	2/8/2017 21:55	SMPL
I1618-09	ME5BR6	7511	1.5451 -	ISM02.3	1	2/8/2017 21:57	SMPL
I1618-10	ME5BR7	9925	2.0369 -	ISM02.3	1	2/8/2017 21:59	SMPL
I1618-11	ME5BR8	3946	0.8188 -	ISM02.3	1	2/8/2017 22:02	SMPL
I1618-12	ME5BR9	9989	2.0499 -	ISM02.3	1	2/8/2017 22:04	SMPL
I1618-13	ME5BS0	4581	0.9482 -	ISM02.3	1	2/8/2017 22:06	SMPL
I1618-14	ME5BS1	3621	0.7526 -	ISM02.3	1	2/8/2017 22:08	SMPL
I1618-15	ME5BS2	3090	0.6445 -	ISM02.3	1	2/8/2017 22:10	SMPL

LB85757

INSTRUMENT ID : CV1

I1618-16	ME5BS2D	3103	0.6471 -	ISM02.3	1 2/8/2017 22:12 SMPL
I1618-17	ME5BS2S	13931	2.853 -	ISM02.3	1 2/8/2017 22:17 SMPL
CCV087	CCV087	23488	4.8 -	ISM02.3	1 2/8/2017 22:19 SMPL
CCB087	CCB087	27	0.0205 -	ISM02.3	1 2/8/2017 22:21 SMPL
I1618-06DLX5	ME5BR3	17962	3.6742 -	ISM02.3	1 2/8/2017 22:23 SMPL
I1618-07DLX5	ME5BR4	12879	2.6387 -	ISM02.3	1 2/8/2017 22:25 SMPL
CCV088	CCV088	23238	4.7491 -	ISM02.3	1 2/8/2017 22:27 SMPL
CCB088	CCB088	28	0.0207 -	ISM02.3	1 2/8/2017 22:30 SMPL