



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB85765
Balance ID #: Gravimetric
Extraction IN Time: 02/10/2017 11:20

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 02/10/2017
Analysis Date: 02/10/2017
Extraction OUT Time: 02/10/2017 11:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight	Weight After Drying	Final Weight After Drying	Change Weight (g)	Result in ppm
1	LB85765BL	LB85765BL	WATER	2	1000	100	3.8542	3.8542	3.8542	3.8542	0	0
2	LB85765BS	LB85765BS	WATER	2	1000	100	2.7411	2.7411	2.7598	2.7598	0.0187	18.7
3	LB85765BSD	LB85765BSD	WATER	2	1000	100	2.8496	2.8496	2.8685	2.8685	0.0189	18.9
4	I1773-01	DSN001	WATER	2	30942	100	3.0942	3.0942	3.0948	3.0948	0.0006	0.02



QC Batch# LB85765 Test: TPH Analysis Date: 02/10/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2203
Magnesium Sulfate	
Sodium Sulfate	EP1722
1:1 HCL	WP50630
Silica Gel	W2061 (3g)
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP52823
LCSWD	10.00 ML	WP52823
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID #

Before Analysis

0.0020 gram Balance: 0.0019 In OVEN TEMP : 70 Dessicator Time In : 13:00
1.0000 gram Balance: 1.0004 In Time: 12:30
Bal Check Time: 11:30

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 14:35
1.0000 gram Balance: 1.0005 Out Time: 14:00
Bal Check Time: 15:20



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB85765
Balance ID # Gravimetric
Extraction IN Time 02/10/2017 11:20

ANALYST: JIGNESH
REVIEWED BY: *rf*
Extraction Date: 02/10/2017
Analysis Date: 02/10/2017
Extraction OUT Time: 02/10/2017 11:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight	Weight After Drying	Final Weight After Drying	Change Weight (g)	Result in ppm
1	LB85765BL	LB85765BL	WATER	2	1000	100	3.8542	3.8542	3.8542	3.8542	0	0
2	LB85765BS	LB85765BS	WATER	2	1000	100	2.7411	2.7411	2.7598	2.7598	0.0187	18.7
3	LB85765BSD	LB85765BSD	WATER	2	1000	100	2.8496	2.8496	2.8685	2.8685	0.0189	18.9
4	I1773-01	DSN001	WATER	2	30942	100	3.0942	3.0942	3.0948	3.0948	0.0006	0.02

used PH strips

→ 0.0-14.0 → W2105

CHEMTECH

QC Batch# LB85765Test: TPHAnalysis Date: 02/10/2017**Chemicals Used:**

Chemical Name	Chemical Lot #
Hexane	W2203
Magnesium Sulfate	
Sodium Sulfate	EP1722
1:1 HCL	WP50630
Silica Gel	W2061 (3g)
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP52823
LCSWD	10.00 ML	WP52823
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator DataAnalytical Balance ID # WCSC-3Before Analysis0.0020 gram Balance: 0.0019In OVEN TEMP : 70Dessicator Time In : 13:001.0000 gram Balance: 1.0004In Time: 12:30Bal Check Time: 11:30After Analysis0.0020 gram Balance: 0.0021Out OVEN TEMP: 71Dessicator Time Out: 14:351.0000 gram Balance: 1.0005Out Time: 14:00Bal Check Time: 15:20used oven #2

LB 85765

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph 11773

WorkList ID : 95509

Date : 2/10/2017 11:17:20 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	11773-01 - 1	TPH	Conc H2SO4 tr	NATI04	H12	DSN001	02/08/2017	1664A

Date/Time 02/10/17 11:17 AM
Received by: 20
Relinquished by: [Signature]

Date/Time 02/10/17 2:30 PM
Received by: [Signature]
Relinquished by: 40