


245.1

LB 85766

Linear

 μ Abs.:

34791



A= 0.0000e+000

B= 2.8795e-004

C= 1.6975e-002

Rho= 0.9998284

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	3	0.000	3				
0.2	0.200	0.236	0.036	760	0.0 %	760				
2.5	2.500	2.520	0.020	8692	0.0 %	8692				
5.0	5.000	4.848	-0.152	16779	0.0 %	16779				
7.5	7.500	7.543	0.043	26137	0.0 %	26137				
10.0	10.000	10.035	0.035	34791	0.0 %	34791				

LB85766 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 Feb 2017 11:43:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	245.1	PPB	10 Feb 2017 12:03:29	S	Std
0.2 - 1	80.2	760	-	0.2000	245.1	PPB	10 Feb 2017 12:05:37	S	Std
2.5 - 1	82.5	8692	-	2.5000	245.1	PPB	10 Feb 2017 12:07:46	S	Std
5.0 - 1	55	16779	-	5.0000	245.1	PPB	10 Feb 2017 12:09:54	S	Std
7.5 - 1	59.5	26137	-	7.5000	245.1	PPB	10 Feb 2017 12:12:01	S	Std
10.0 - 1	310	34791	-	10.0000	245.1	PPB	10 Feb 2017 12:14:09	S	Std
ICV24 - 1	ICV24	13480	3.8985	-245.1		PPB	10 Feb 2017 12:16:54	U	SMPL
ICB24 - 1	ICB24	-23	0.0104	-245.1		PPB	10 Feb 2017 12:19:01	U	SMPL
CCV93 - 1	CCV93	16715	4.8300	-245.1		PPB	10 Feb 2017 12:21:07	U	SMPL
CCB93 - 1	CCB93	-19	0.0115	-245.1		PPB	10 Feb 2017 12:23:14	U	SMPL
CRA - 1	CRA	736	0.2289	-245.1		PPB	10 Feb 2017 12:25:20	U	SMPL
HighStd - 1	HighStd	34002	9.8078	-245.1		PPB	10 Feb 2017 12:27:26	U	SMPL
ChkStd - 1	ChkStd	23157	6.6850	-245.1		PPB	10 Feb 2017 12:29:33	U	SMPL
PB96840BL - 1	PBW	-53	0.0017	-245.1		PPB	10 Feb 2017 12:31:41	U	SMPL
PB96840BS - 1	LCSW	12308	3.5610	-245.1		PPB	10 Feb 2017 12:33:49	U	SMPL
I1739-01 - 1	020217-PR-E-D-M	44	0.0296	-245.1		PPB	10 Feb 2017 12:35:56	U	SMPL
I1739-02 - 1	020217-PR-E-U-M	50	0.0314	-245.1		PPB	10 Feb 2017 12:40:28	U	SMPL
I1740-01 - 1	020317-PR-E-D-M	49	0.0311	-245.1		PPB	10 Feb 2017 12:42:36	U	SMPL
I1740-02 - 1	020317-PR-E-U-M	45	0.0299	-245.1		PPB	10 Feb 2017 12:44:45	U	SMPL
I1741-01 - 1	020617-PR-E-D-M	59	0.0340	-245.1		PPB	10 Feb 2017 12:46:52	U	SMPL
CCV94 - 1	CCV94	16529	4.7765	-245.1		PPB	10 Feb 2017 12:48:58	U	SMPL
CCB94 - 1	CCB94	-17	0.0121	-245.1		PPB	10 Feb 2017 12:51:05	U	SMPL
I1741-02 - 1	020617-PR-E-U-M	67	0.0363	-245.1		PPB	10 Feb 2017 12:53:12	U	SMPL
I1766-01 - 1	020717-PR-E-D-M	49	0.0311	-245.1		PPB	10 Feb 2017 12:55:18	U	SMPL
I1766-02 - 1	020717-PR-E-U-M	42	0.0291	-245.1		PPB	10 Feb 2017 12:57:25	U	SMPL
I1767-01 - 1	020817-PRRT-F-D-M	47	0.0305	-245.1		PPB	10 Feb 2017 12:59:32	U	SMPL
I1767-02 - 1	020817-PRRT-F-U-M	102	0.0463	-245.1		PPB	10 Feb 2017 13:01:40	U	SMPL
I1767-02DUP - 1	020817-PRRT-F-U-MDUP	95	0.0443	-245.1		PPB	10 Feb 2017 13:03:48	U	SMPL
I1767-02MS - 1	020817-PRRT-F-U-MMS	14081	4.0716	-245.1		PPB	10 Feb 2017 13:05:56	U	SMPL
I1767-02MSD - 1	020817-PRRT-F-U-MMSD	14442	4.1755	-245.1		PPB	10 Feb 2017 13:08:04	U	SMPL
I1767-02LX5 - 1		-13	0.0132	-245.1		PPB	10 Feb 2017 13:10:13	U	SMPL
I1767-02A - 1		14140	4.0886	-245.1		PPB	10 Feb 2017 13:12:20	U	SMPL
CCV95 - 1	CCV95	16489	4.7650	-245.1		PPB	10 Feb 2017 13:14:26	U	SMPL
CCB95 - 1	CCB95	-37	0.0063	-245.1		PPB	10 Feb 2017 13:16:33	U	SMPL