

LB85768



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 300 Nitrate+Nitrite
Parameter: Nitrate+Nitrite
Run Number: LB85768
Instrument: IC Metrohm

M300.0 - Inorganic Nitrogen-16



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analytical Summary Report

Analysis Method: 300 Nitrate+Nitrite
Parameter: Nitrate+Nitrite
Run Number: LB85768
Instrument: IC Metrohm

Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb	Dil	Analytical Date
1	STD1	STD1	CAL	0.000	1	1/16/17
2	STD2	STD2	CAL	0.685	1	1/16/17
3	STD3	STD3	CAL	1.327	1	1/16/17
4	STD4	STD4	CAL	2.212	1	1/16/17
5	STD5	STD5	CAL	5.053	1	1/16/17
6	STD6	STD6	CAL	10.162	1	1/16/17
7	STD7	STD7	CAL	12.735	1	1/16/17
8	STD8	STD8	CAL	26.936	1	1/16/17
9	ICV	ICV	ICV	13.761	1	1/16/17
10	ICB	ICB	ICB	0.000	1	1/16/17
11	CCV	CCV	CCV	12.677	1	2/8/17
12	CCB	CCB	CCB	0.000	1	2/8/17
13	LB85768BLW	LB85768BLW	MB	0.000	1	2/8/17
14	LB85768BSW	LB85768BSW	LCS	13.464	1	2/8/17
15	I1773-01	DSN001	SAM	0.782	1	2/8/17
16	I1773-01D	DSN001D	DUP	0.785	1	2/8/17
17	I1773-01S	DSN001S	MS	12.177	1	2/8/17
18	CCV	CCV	CCV	13.203	1	2/8/17
19	CCB	CCB	CCB	0.000	1	2/8/17

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/10/17


Analyst : AP

Data File : LB85768.MDB

Approved By :

Approved Date :

Worksheet # :


2/10/17

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Nitrate+Nitrite												
STD1	STD1				W	1/16/17						
Nitrate+Nitrite		PASS	0.000		0.000						mg/L	
STD2	STD2				W	1/16/17						
Nitrate+Nitrite		PASS	0.685		0.690						mg/L	
STD3	STD3				W	1/16/17						
Nitrate+Nitrite		PASS	1.327		1.33						mg/L	
STD4	STD4				W	1/16/17						
Nitrate+Nitrite		PASS	2.212		2.21						mg/L	
STD5	STD5				W	1/16/17						
Nitrate+Nitrite		PASS	5.053		5.05						mg/L	
STD6	STD6				W	1/16/17						
Nitrate+Nitrite		PASS	10.162		10.20						mg/L	
STD7	STD7				W	1/16/17						
Nitrate+Nitrite		PASS	12.735		12.70						mg/L	
STD8	STD8				W	1/16/17						
Nitrate+Nitrite		PASS	26.936		26.90						mg/L	
ICV1	ICV1				W	1/16/17						
Nitrate+Nitrite		PASS	13.761		13.80	100.0	90	110			mg/L	
ICB1	ICB1				W	1/16/17						
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500			mg/L	
CCV1	CCV1				W	2/8/17						
Nitrate+Nitrite		PASS	12.677		12.70	96.0	90	110			mg/L	
CCB1	CCB1				W	2/8/17						
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500			mg/L	
LB85768BLW	LB85768BLW				W	2/8/17						
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500			mg/L	
LB85768BSW	LB85768BSW				W	2/8/17						
Nitrate+Nitrite		PASS	13.464		13.50	98.0	90	110			mg/L	
I1773-01	DSN001		1		W	2/8/17						
Nitrate+Nitrite		PASS	0.782		0.780						mg/L	
I1773-01D	DSN001D		1		W	2/8/17						
Nitrate+Nitrite		PASS	0.785		0.79				1.3	20	mg/L	
I1773-01S	DSN001S		1		W	2/8/17						
Nitrate+Nitrite		PASS	12.177		12.20	87.0	80	120			mg/L	
CCV2	CCV2				W	2/8/17						
Nitrate+Nitrite		PASS	13.203		13.20	100.0	90	110			mg/L	
CCB2	CCB2				W	2/8/17						
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500			mg/L	