


7471A

LB 85770  
INSTRUMENT ID: CV 4

Linear

 $\mu$  Abs.:

41392



A= 0.0000e+000

B= 2.4265e-004

C= -3.8075e-002

Rho= 0.9999595

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | -0.027 | -0.027 | 44    | 0.000      | 44    |       |       |       |       |
| 0.2    | 0.200  | 0.191  | -0.009 | 945   | 0.0 %      | 945   |       |       |       |       |
| 2.5    | 2.500  | 2.570  | 0.070  | 10749 | 0.0 %      | 10749 |       |       |       |       |
| 5.0    | 5.000  | 4.976  | -0.024 | 20663 | 0.0 %      | 20663 |       |       |       |       |
| 7.5    | 7.500  | 7.484  | -0.016 | 31001 | 0.0 %      | 31001 |       |       |       |       |
| 10.0   | 10.000 | 10.006 | 0.006  | 41392 | 0.0 %      | 41392 |       |       |       |       |

## LB85770 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 10 Feb 2017 13:53:31

| Sample ID        | Extended ID        | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|------------------|--------------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1          | 50                 | 44         | -       | 0.0000   | 7471A  | PPB   | 10 Feb 2017 13:55:20 | S    | Std  |
| 0.2 - 1          | 50.2               | 945        | -       | 0.2000   | 7471A  | PPB   | 10 Feb 2017 13:57:37 | S    | Std  |
| 2.5 - 1          | 22.5               | 10749      | -       | 2.5000   | 7471A  | PPB   | 10 Feb 2017 13:59:54 | S    | Std  |
| 5.0 - 1          | 55                 | 20663      | -       | 5.0000   | 7471A  | PPB   | 10 Feb 2017 14:02:12 | S    | Std  |
| 7.5 - 1          | 57.5               | 31001      | -       | 7.5000   | 7471A  | PPB   | 10 Feb 2017 14:04:29 | S    | Std  |
| 10.0 - 1         | 51.0               | 41392      | -       | 10.0000  | 7471A  | PPB   | 10 Feb 2017 14:06:46 | S    | Std  |
| ICV25 - 1        | ICV25              | 17441      | 4.1940  | -        | 7471A  | PPB   | 10 Feb 2017 14:09:37 | U    | SMPL |
| ICB25 - 1        | ICB25              | 9          | -0.0359 | -        | 7471A  | PPB   | 10 Feb 2017 14:11:52 | U    | SMPL |
| CCV96 - 1        | CCV96              | 20514      | 4.9397  | -        | 7471A  | PPB   | 10 Feb 2017 14:14:08 | U    | SMPL |
| CCB96 - 1        | CCB96              | -7         | -0.0398 | -        | 7471A  | PPB   | 10 Feb 2017 14:16:24 | U    | SMPL |
| CRA - 1          | CRA                | 926        | 0.1866  | -        | 7471A  | PPB   | 10 Feb 2017 14:18:40 | U    | SMPL |
| HighStd - 1      | HighStd            | 41047      | 9.9221  | -        | 7471A  | PPB   | 10 Feb 2017 14:20:55 | U    | SMPL |
| ChkStd - 1       | ChkStd             | 31293      | 7.5552  | -        | 7471A  | PPB   | 10 Feb 2017 14:23:12 | U    | SMPL |
| PB96861BL - 1    | PBS                | -8         | -0.0400 | -        | 7471A  | PPB   | 10 Feb 2017 14:25:28 | U    | SMPL |
| PB96861BS - 1    | LCSS               | 13900      | 3.3348  | -        | 7471A  | PPB   | 10 Feb 2017 14:27:45 | U    | SMPL |
| I1745-03 - 1     | SB-1               | 65045      | 15.7452 | -        | 7471A  | PPB   | 10 Feb 2017 14:30:02 | U    | SMPL |
| I1751-01 - 1     | ROLL-OFF-COMP      | 3711       | 0.8624  | -        | 7471A  | PPB   | 10 Feb 2017 14:32:19 | U    | SMPL |
| I1751-01DUP - 1  | ROLL-OFF-COMPDUP   | 3210       | 0.7408  | -        | 7471A  | PPB   | 10 Feb 2017 14:34:37 | U    | SMPL |
| I1751-01MS - 1   | ROLL-OFF-COMPMS    | 17065      | 4.1028  | -        | 7471A  | PPB   | 10 Feb 2017 14:36:54 | U    | SMPL |
| I1751-01MSD - 1  | ROLL-OFF-COMPMSD   | 17600      | 4.2326  | -        | 7471A  | PPB   | 10 Feb 2017 14:39:10 | U    | SMPL |
| CCV97 - 1        | CCV97              | 20541      | 4.9462  | -        | 7471A  | PPB   | 10 Feb 2017 14:41:26 | U    | SMPL |
| CCB97 - 1        | CCB97              | 0          | -0.0381 | -        | 7471A  | PPB   | 10 Feb 2017 14:43:42 | U    | SMPL |
| I1754-09 - 1     | E2-BI-B21-CONCRETE | 890        | 0.1779  | -        | 7471A  | PPB   | 10 Feb 2017 14:45:58 | U    | SMPL |
| I1751-01LX5 - 1  |                    | 793        | 0.1543  | -        | 7471A  | PPB   | 10 Feb 2017 14:48:15 | U    | SMPL |
| I1751-01A - 1    |                    | 21027      | 5.0642  | -        | 7471A  | PPB   | 10 Feb 2017 14:50:31 | U    | SMPL |
| I1745-03DLX5 - 1 |                    | 12792      | 3.0659  | -        | 7471A  | PPB   | 10 Feb 2017 14:52:47 | U    | SMPL |
| CCV98 - 1        | CCV98              | 20479      | 4.9312  | -        | 7471A  | PPB   | 10 Feb 2017 14:55:04 | U    | SMPL |
| CCB98 - 1        | CCB98              | 0          | -0.0381 | -        | 7471A  | PPB   | 10 Feb 2017 14:57:21 | U    | SMPL |