

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/13/17

Analyst : AK

Data File : LB85772.csv

Approved By : 

Approved Date : 2/13/17

Worksheet # :

m9012 A-13 Total Amenable & Reactive cyanide -16

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB	Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units
										Line 2
Cyanide										
0.0PPBCN	0.0PPBCN			W	2/10/17					
Cyanide	PASS	1.832	0.002							mg/L
5.0PPBCN	5.0PPBCN			W	2/10/17					
Cyanide	PASS	4.706	0.005							mg/L
10PPBCN	10PPBCN			W	2/10/17					
Cyanide	PASS	9.625	0.010							mg/L
50PPBCN	50PPBCN			W	2/10/17					
Cyanide	PASS	47.966	0.048							mg/L
100PPBCN	100PPBCN			W	2/10/17					
Cyanide	PASS	100.528	0.101							mg/L
250PPBCN	250PPBCN			W	2/10/17					
Cyanide	PASS	250.470	0.250							mg/L
500PPBCN	500PPBCN			W	2/10/17					
Cyanide	PASS	499.873	0.500							mg/L
ICV1	ICV1			W	2/10/17					
Cyanide	PASS	93.914	0.09		90.0	90	110			mg/L
ICB1	ICB1			W	2/10/17					
Cyanide	PASS	0.475	0.000				+/-0.0030			mg/L
CCV1	CCV1			W	2/10/17					
Cyanide	PASS	235.508	0.24		96.0	90	110			mg/L
CCB1	CCB1			W	2/10/17					
Cyanide	PASS	0.336	0.000				+/-0.0030			mg/L
PB96862BLS	PB96862BLS			S	2/10/17					
Cyanide	PASS	0.270	0.014				+/-0.0330			mg/Kg
PB96862BSS	PB96862BSS			S	2/10/17					
Cyanide	PASS	96.822	4.84		97.0	85.00	115.00			mg/Kg
LOWPB96862	LOWPB96862			W	2/10/17					
Cyanide	PASS	10.344	0.010							mg/L
HIGHPB96862	HIGHPB96862			W	2/10/17					
Cyanide	PASS	488.645	0.489							mg/L
I1777-04	38-SB-03-0-2		1	S	2/10/17					
Cyanide	PASS	0.640	0.034							mg/Kg
I1777-05	38-SB-102-0-2		1	S	2/10/17					
Cyanide	PASS	1.087	0.060							mg/Kg
CCV2	CCV2			W	2/10/17					
Cyanide	PASS	242.006	0.24		96.0	90	110			mg/L
CCB2	CCB2			W	2/10/17					
Cyanide	PASS	0.458	0.000				+/-0.0030			mg/L
I1777-06	38-SB-03-11-13		1	S	2/10/17					
Cyanide	PASS	0.399	0.021							mg/Kg

Chemtech Consulting Group


Analytical Review Report

Date Printed : 2/13/17

Analyst : AK

Data File : LB85772.csv

Approved By : 

Approved Date : 

Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2	
Cyanide													
I1777-07	38-SB-02-5-7			1	S	2/10/17							
Cyanide		PASS	0.142		0.008						mg/Kg		
I1777-08	38-SB-01-0-2			1	S	2/10/17							
Cyanide		PASS	1.079		0.061						mg/Kg		
I1777-01	38-SB-02-0-2			1	S	2/10/17							
Cyanide		PASS	1.028		0.056						mg/Kg		
I1777-01D	38-SB-02-0-2D			1	S	2/10/17							
Cyanide		FAIL	1.595		0.087				43.4	20	mg/Kg		
I1777-02S	38-SB-02-0-2S			1	S	2/10/17							
Cyanide		PASS	39.728		2.140	96.0	75	125			mg/Kg		
I1777-03SD	38-SB-02-0-2SD			1	S	2/10/17							
Cyanide		PASS	40.716		2.180	99.0	75	125	1.9	20	mg/Kg		
CCV3	CCV3				W	2/10/17							
Cyanide		PASS	258.531		0.26	104.0	90	110			mg/L		
CCB3	CCB3				W	2/10/17							
Cyanide		PASS	0.096		0.000			+/-0.0030			mg/L		