


7470A

LB 85784
INSTRUMENT ID: CV4

Linear

 μ Abs.:

36815



A= 0.0000e+000

B= 2.7253e-004

C= 1.6951e-003

Rho= 0.9999159

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	40	0.000	40				
0.2	0.200	0.217	0.017	790	0.0 %	790				
2.5	2.500	2.529	0.029	9274	0.0 %	9274				
5.0	5.000	4.895	-0.105	17954	0.0 %	17954				
7.5	7.500	7.512	0.012	27557	0.0 %	27557				
10.0	10.000	10.035	0.035	36815	0.0 %	36815				

LB85784 INMSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Feb 2017 13:56:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	40	-	0.0000	7470A	PPB	11 Feb 2017 14:03:51	S	Std
0.2 - 1	SD-2	790	-	0.2000	7470A	PPB	11 Feb 2017 14:05:59	S	Std
2.5 - 1	SD-5	9274	-	2.5000	7470A	PPB	11 Feb 2017 14:08:08	S	Std
5.0 - 1	SD	17954	-	5.0000	7470A	PPB	11 Feb 2017 14:10:16	S	Std
7.5 - 1	SD	27557	-	7.5000	7470A	PPB	11 Feb 2017 14:12:24	S	Std
10.0 - 1	SD	36815	-	10.0000	7470A	PPB	11 Feb 2017 14:14:31	S	Std
ICV26 - 1	ICV26	13679	3.7296	-	7470A	PPB	11 Feb 2017 14:17:04	U	SMPL
ICB26 - 1	ICB26	27	0.0091	-	7470A	PPB	11 Feb 2017 14:19:10	U	SMPL
CCV03 - 1	CCV03	17604	4.7993	-	7470A	PPB	11 Feb 2017 14:21:16	U	SMPL
CCB03 - 1	CCB03	17	0.0063	-	7470A	PPB	11 Feb 2017 14:23:23	U	SMPL
CRA - 1	CRA	768	0.2110	-	7470A	PPB	11 Feb 2017 14:25:29	U	SMPL
HighStd - 1	HighStd	36221	9.8730	-	7470A	PPB	11 Feb 2017 14:27:36	U	SMPL
ChkStd - 1	ChkStd	25602	6.9790	-	7470A	PPB	11 Feb 2017 14:29:43	U	SMPL
PB96876BL - 1	PBW	41	0.0129	-	7470A	PPB	11 Feb 2017 14:31:50	U	SMPL
PB96876BS - 1	LCSW	13475	3.6740	-	7470A	PPB	11 Feb 2017 14:33:57	U	SMPL
PB96876TB - 1	PB96849TB	91	0.0265	-	7470A	PPB	11 Feb 2017 14:36:05	U	SMPL
I1769-02 - 1	9717	108	0.0311	-	7470A	PPB	11 Feb 2017 14:38:13	U	SMPL
I1769-02DUP - 1	9717DUP	100	0.0289	-	7470A	PPB	11 Feb 2017 14:40:22	U	SMPL
I1769-02MS - 1	9717MS	12921	3.5230	-	7470A	PPB	11 Feb 2017 14:42:30	U	SMPL
I1769-02MSD - 1	9717MSD	13351	3.6402	-	7470A	PPB	11 Feb 2017 14:44:37	U	SMPL
CCV04 - 1	CCV04	17706	4.8271	-	7470A	PPB	11 Feb 2017 14:46:44	U	SMPL
CCB04 - 1	CCB04	16	0.0061	-	7470A	PPB	11 Feb 2017 14:48:51	U	SMPL
I1769-02LX5 - 1		56	0.0170	-	7470A	PPB	11 Feb 2017 14:50:58	U	SMPL
I1769-02A - 1		13598	3.7075	-	7470A	PPB	11 Feb 2017 14:53:05	U	SMPL
CCV05 - 1	CCV05	17630	4.8064	-	7470A	PPB	11 Feb 2017 14:55:11	U	SMPL
CCB05 - 1	CCB05	4	0.0028	-	7470A	PPB	11 Feb 2017 14:57:18	U	SMPL