


7471A

LAB 5785
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47070



A= 0.0000e+000

B= 2.1311e-004

C= -2.9385e-002

Rho= 0.9999610

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	39	0.000	39				
0.2	0.200	0.207	0.007	1107	0.0 %	1107				
2.5	2.500	2.493	-0.007	11834	0.0 %	11834				
5.0	5.000	5.063	0.063	23897	0.0 %	23897				
7.5	7.500	7.457	-0.043	35131	0.0 %	35131				
10.0	10.000	10.002	0.002	47070	0.0 %	47070				

LB85785 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 11 Feb 2017 15:38:06

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7471A	PPB	11 Feb 2017 16:30:13	S	Std
0.2 - 1	50.2	1107	-	0.2000	7471A	PPB	11 Feb 2017 16:32:30	S	Std
2.5 - 1	52.5	11834	-	2.5000	7471A	PPB	11 Feb 2017 16:34:48	S	Std
5.0 - 1	55	23897	-	5.0000	7471A	PPB	11 Feb 2017 16:37:05	S	Std
7.5 - 1	57.5	35131	-	7.5000	7471A	PPB	11 Feb 2017 16:39:23	S	Std
10.0 - 1	510	47070	-	10.0000	7471A	PPB	11 Feb 2017 16:46:46	S	Std
ICV27 - 1	ICV27	18791	3.9751	-7471A	PPB	11 Feb 2017 16:49:29	U	SMPL	
ICB27 - 1	ICB27	73	-0.0138	-7471A	PPB	11 Feb 2017 16:51:44	U	SMPL	
CCV06 - 1	CCV06	24008	5.0869	-7471A	PPB	11 Feb 2017 16:54:00	U	SMPL	
CCB06 - 1	CCB06	81	-0.0121	-7471A	PPB	11 Feb 2017 16:56:15	U	SMPL	
CRA - 1	CRA	1118	0.2089	-7471A	PPB	11 Feb 2017 16:58:31	U	SMPL	
HighStd - 1	HighStd	47657	10.1266	-7471A	PPB	11 Feb 2017 17:00:47	U	SMPL	
ChkStd - 1	ChkStd	33120	7.0287	-7471A	PPB	11 Feb 2017 17:03:03	U	SMPL	
PB96877BL - 1	PBS	29	-0.0232	-7471A	PPB	11 Feb 2017 17:05:19	U	SMPL	
PB96877BS - 1	LCSS	17507	3.7015	-7471A	PPB	11 Feb 2017 17:07:36	U	SMPL	
I1303-01 - 1	0022	194190	41.3538	-7471A	PPB	11 Feb 2017 17:09:53	U	SMPL	
CCV07 - 1	CCV07	23820	5.0468	-7471A	PPB	11 Feb 2017 17:16:49	U	SMPL	
CCB07 - 1	CCB07	58	-0.0170	-7471A	PPB	11 Feb 2017 17:19:07	U	SMPL	
I1303-01DLX10 - 1	0022	19317	4.0872	-7471A	PPB	11 Feb 2017 17:23:39	U	SMPL	
CCV08 - 1	CCV08	23912	5.0664	-7471A	PPB	11 Feb 2017 17:28:11	U	SMPL	
CCB08 - 1	CCB08	91	-0.0100	-7471A	PPB	11 Feb 2017 17:30:26	U	SMPL	