


7471A

LB85786
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47183



A= 0.0000e+000

B= 2.1423e-004

C= -5.2953e-002

Rho= 0.9998874

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	124	0.000	124				
0.2	0.200	0.220	0.020	1274	0.0 %	1274				
2.5	2.500	2.578	0.078	12281	0.0 %	12281				
5.0	5.000	4.927	-0.073	23247	0.0 %	23247				
7.5	7.500	7.446	-0.054	35005	0.0 %	35005				
10.0	10.000	10.055	0.055	47183	0.0 %	47183				

LB85786 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 12 Feb 2017 12:45:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	124	-	0.0000	7471A	PPB	12 Feb 2017 12:48:05	S	Std
0.2 - 1	50.2	1274	-	0.2000	7471A	PPB	12 Feb 2017 12:50:23	S	Std
2.5 - 1	52.5	12281	-	2.5000	7471A	PPB	12 Feb 2017 12:52:40	S	Std
5.0 - 1	55	23247	-	5.0000	7471A	PPB	12 Feb 2017 12:54:58	S	Std
7.5 - 1	57.5	35005	-	7.5000	7471A	PPB	12 Feb 2017 13:02:16	S	Std
10.0 - 1	510	47183	-	10.0000	7471A	PPB	12 Feb 2017 13:07:08	S	Std
ICV28 - 1	ICV28	18971	4.0112	-	7471A	PPB	12 Feb 2017 13:09:48	U	SMPL
ICB28 - 1	ICB28	61	-0.0399	-	7471A	PPB	12 Feb 2017 13:12:03	U	SMPL
CCV09 - 1	CCV09	22991	4.8724	-	7471A	PPB	12 Feb 2017 13:14:19	U	SMPL
CCB09 - 1	CCB09	50	-0.0422	-	7471A	PPB	12 Feb 2017 13:16:34	U	SMPL
CRA - 1	CRA	1288	0.2230	-	7471A	PPB	12 Feb 2017 13:18:50	U	SMPL
HighStd - 1	HighStd	47817	10.1909	-	7471A	PPB	12 Feb 2017 13:21:06	U	SMPL
ChkStd - 1	ChkStd	31394	6.6726	-	7471A	PPB	12 Feb 2017 13:23:22	U	SMPL
PB96880BL - 1	PBS	235	-0.0026	-	7471A	PPB	12 Feb 2017 13:25:38	U	SMPL
PB96880BS - 1	LCSS	16070	3.3897	-	7471A	PPB	12 Feb 2017 13:27:54	U	SMPL
I1777-01 - 1	38-SB-02-0-2	11698	2.4531	-	7471A	PPB	12 Feb 2017 13:35:04	U	SMPL
I1777-01DUP - 1	38-SB-02-0-2DUP	11020	2.3079	-	7471A	PPB	12 Feb 2017 13:37:21	U	SMPL
I1777-02 - 1	I1777-01MS	27405	5.8180	-	7471A	PPB	12 Feb 2017 13:42:30	U	SMPL
I1777-03 - 1	I1777-01MSD	27606	5.8611	-	7471A	PPB	12 Feb 2017 13:44:47	U	SMPL
I1777-04 - 1	38-SB-03-0-2	3968	0.7971	-	7471A	PPB	12 Feb 2017 13:47:03	U	SMPL
CCV10 - 1	CCV10	23181	4.9131	-	7471A	PPB	12 Feb 2017 13:49:19	U	SMPL
CCB10 - 1	CCB10	74	-0.0371	-	7471A	PPB	12 Feb 2017 13:51:35	U	SMPL
I1777-05 - 1	38-SB-102-0-2	11234	2.3537	-	7471A	PPB	12 Feb 2017 13:53:51	U	SMPL
I1777-06 - 1	38-SB-03-11-13	425	0.0381	-	7471A	PPB	12 Feb 2017 13:56:06	U	SMPL
I1777-07 - 1	38-SB-02-5-7	1153	0.1941	-	7471A	PPB	12 Feb 2017 13:58:22	U	SMPL
I1777-08 - 1	38-SB-01-0-2	5873	1.2052	-	7471A	PPB	12 Feb 2017 14:00:38	U	SMPL
I1777-01LX5 - 1		2550	0.4933	-	7471A	PPB	12 Feb 2017 14:02:55	U	SMPL
I1777-01A - 1		29758	6.3221	-	7471A	PPB	12 Feb 2017 14:05:11	U	SMPL
CCV11 - 1	CCV11	22834	4.8388	-	7471A	PPB	12 Feb 2017 14:07:29	U	SMPL
CCB11 - 1	CCB11	70	-0.0380	-	7471A	PPB	12 Feb 2017 14:09:46	U	SMPL