


245.1

LB85787
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37526



A= 0.0000e+000

B= 2.7103e-004

C= -1.0502e-002

Rho= 0.9994796

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	103	0.000	103				
0.2	0.200	0.220	0.020	849	0.0 %	849				
2.5	2.500	2.544	0.044	9424	0.0 %	9424				
5.0	5.000	4.997	-0.003	18477	0.0 %	18477				
7.5	7.500	7.262	-0.238	26832	0.0 %	26832				
10.0	10.000	10.160	0.160	37526	0.0 %	37526				

LB85787 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Feb 2017 14:58:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	103	-	0.0000	245.1	PPB	12 Feb 2017 15:04:46	S	Std
0.2 - 1	S0.2	849	-	0.2000	245.1	PPB	12 Feb 2017 15:06:54	S	Std
2.5 - 1	S2.5	9424	-	2.5000	245.1	PPB	12 Feb 2017 15:09:02	S	Std
5.0 - 1	S5	18477	-	5.0000	245.1	PPB	12 Feb 2017 15:11:10	S	Std
7.5 - 1	S7.5	26832	-	7.5000	245.1	PPB	12 Feb 2017 15:13:18	S	Std
10.0 - 1	S10	37526	-	10.0000	245.1	PPB	12 Feb 2017 15:20:17	S	Std
ICV29 - 1	ICV29	14241	3.8492	-245.1		PPB	12 Feb 2017 15:23:00	U	SMPL
ICB29 - 1	ICB29	65	0.0071	-245.1		PPB	12 Feb 2017 15:25:06	U	SMPL
CCV12 - 1	CCV12	18375	4.9697	-245.1		PPB	12 Feb 2017 15:27:13	U	SMPL
CCB12 - 1	CCB12	68	0.0079	-245.1		PPB	12 Feb 2017 15:29:20	U	SMPL
CRA - 1	CRA	861	0.2229	-245.1		PPB	12 Feb 2017 15:31:26	U	SMPL
HighStd - 1	HighStd	36545	9.8943	-245.1		PPB	12 Feb 2017 15:33:33	U	SMPL
ChkStd - 1	ChkStd	23512	6.3620	-245.1		PPB	12 Feb 2017 15:35:40	U	SMPL
PB96878BL - 1	PBW	70	0.0085	-245.1		PPB	12 Feb 2017 15:37:48	U	SMPL
PB96878BS - 1	LCSW	14534	3.9287	-245.1		PPB	12 Feb 2017 15:39:56	U	SMPL
I1299-04 - 1	0011B	48785	13.2117	-245.1		PPB	12 Feb 2017 15:44:12	U	SMPL
CCV13 - 1	CCV13	18248	4.9353	-245.1		PPB	12 Feb 2017 15:50:37	U	SMPL
CCB13 - 1	CCB13	53	0.0039	-245.1		PPB	12 Feb 2017 15:52:44	U	SMPL
I1299-04DLX2 - 1	0011B	24207	6.5503	-245.1		PPB	12 Feb 2017 15:56:58	U	SMPL
CCV14 - 1	CCV14	18021	4.8737	-245.1		PPB	12 Feb 2017 16:01:11	U	SMPL
CCB14 - 1	CCB14	67	0.0077	-245.1		PPB	12 Feb 2017 16:03:18	U	SMPL