


7470A

LB55788
INSTRUMENT ID: CV4

Linear

 μ Abs.:

36386



A= 0.0000e+000

B= 2.7488e-004

C= -4.1238e-002

Rho= 0.9999536

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	81	0.000	81				
0.2	0.200	0.176	-0.024	791	0.0 %	791				
2.5	2.500	2.510	0.010	9280	0.0 %	9280				
5.0	5.000	5.071	0.071	18597	0.0 %	18597				
7.5	7.500	7.502	0.002	27440	0.0 %	27440				
10.0	10.000	9.961	-0.039	36386	0.0 %	36386				

LB85788 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Feb 2017 16:14:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	81	-	0.0000	7470A	PPB	12 Feb 2017 16:17:44	S	Std
0.2 - 1	50.2	791	-	0.2000	7470A	PPB	12 Feb 2017 16:19:52	S	Std
2.5 - 1	52.5	9280	-	2.5000	7470A	PPB	12 Feb 2017 16:22:01	S	Std
5.0 - 1	55	18597	-	5.0000	7470A	PPB	12 Feb 2017 16:24:09	S	Std
7.5 - 1	57.5	27440	-	7.5000	7470A	PPB	12 Feb 2017 16:26:17	S	Std
10.0 - 1	510	36386	-	10.0000	7470A	PPB	12 Feb 2017 16:28:24	S	Std
ICV30 - 1	ICV30	14699	3.9993	-	7470A	PPB	12 Feb 2017 16:31:28	U	SMPL
ICB30 - 1	ICB30	68	-0.0225	-	7470A	PPB	12 Feb 2017 16:33:34	U	SMPL
CCV15 - 1	CCV15	18404	5.0177	-	7470A	PPB	12 Feb 2017 16:35:41	U	SMPL
CCB15 - 1	CCB15	56	-0.0258	-	7470A	PPB	12 Feb 2017 16:37:47	U	SMPL
CRA - 1	CRA	824	0.1853	-	7470A	PPB	12 Feb 2017 16:39:54	U	SMPL
HighStd - 1	HighStd	36536	10.0019	-	7470A	PPB	12 Feb 2017 16:42:00	U	SMPL
ChkStd - 1	ChkStd	24624	6.7275	-	7470A	PPB	12 Feb 2017 16:44:07	U	SMPL
PB96879BL - 1	PBW	36	-0.0313	-	7470A	PPB	12 Feb 2017 16:46:15	U	SMPL
PB96879BS - 1	LCSW	14452	3.9314	-	7470A	PPB	12 Feb 2017 16:48:23	U	SMPL
I1303-05 - 1	0011B	48411	13.2662	-	7470A	PPB	12 Feb 2017 16:50:30	U	SMPL
CCV16 - 1	CCV16	18605	5.0730	-	7470A	PPB	12 Feb 2017 16:59:02	U	SMPL
CCB16 - 1	CCB16	49	-0.0278	-	7470A	PPB	12 Feb 2017 17:01:09	U	SMPL
I1303-05DLX2 - 1	0011B	23936	6.5384	-	7470A	PPB	12 Feb 2017 17:05:23	U	SMPL
CCV17 - 1	CCV17	18389	5.0136	-	7470A	PPB	12 Feb 2017 17:09:37	U	SMPL
CCB17 - 1	CCB17	59	-0.0250	-	7470A	PPB	12 Feb 2017 17:11:44	U	SMPL