



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB85794

Temp.: 20.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/13/2017	11:00
2	CAL2	1	Water			7.00	02/13/2017	11:01
3	CAL3	1	Water			10.03	02/13/2017	11:02
4	ICV	1	Water			7.00	02/13/2017	11:03
5	CCV1	1	Water			2.02	02/13/2017	11:05
6	I1795-01	1	Water			4.24	02/13/2017	11:06
7	I1795-01DUP	1	Water			4.25	02/13/2017	11:07
8	CCV2	1	Water			12.01	02/13/2017	11:08

21385794

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i1795

WorkList ID : 95562

Date : 2/13/2017 10:49:13 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1795-01	pH	NONE	STER08	H22	POOLWATER	02/10/2017	SM 4500-PH B

Date/Time 02-13-17 10:49 AM
Received by: JP
Relinquished by: JP

Date/Time 02-13-17 11:13 AM
Received by: JP
Relinquished by: JP