



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85797

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2189
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/13/2017	13:35
2	CAL2	1	Water			7.00	02/13/2017	13:36
3	CAL3	1	Water			10.04	02/13/2017	13:37
4	ICV	1	Water			7.00	02/13/2017	13:38
5	CCV1	1	Water			2.02	02/13/2017	13:39
6	I1769-02	1	Solid	20.02	20	10.62	02/13/2017	13:40
7	I1790-02	1	Solid	20.04	20	8.17	02/13/2017	13:41
8	I1790-02DUP	1	Solid	20.02	20	8.18	02/13/2017	13:42
9	CCV2	1	Water			12.01	02/13/2017	13:43

WORKLIST(Hardcopy Internal Chain)

213 85 797

WorkList Name : CORROSIVITY I1790

WorkList ID : 95565

Date : 2/13/2017 1:21:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1769-02	Corrosivity	Cool 4 deg C	SEAC01	H11	9717	02/08/2017	9045C
	Solid	I1790-02	Corrosivity	Cool 4 deg C	PARS15		38-WCS021017-0-15	02/10/2017	9045C

Date/Time 02-13-17 1:21 PM
 Received by: SP
 Relinquished by: SP

Date/Time 02-13-17 5:00 PM
 Received by: SP
 Relinquished by: SP