

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB85801

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP50397
calibration std. phosphate 1 ppm	WP52872
calibration std. phosphate 0.5 ppm	WP52871
calibration std. phosphate 0.3 ppm	WP52870
calibration std. 0 ppm	WP52867
phosphate ICV std.	WP52876
phosphate CCV std.	WP52875
calibration std. phosphate 0.1 ppm	WP52869
calibration std. phosphate 0.05 ppm	WP52868
Combined reagent	WP52874
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP52873

Intercept: -0.0003Slope: 0.6537Regression: 0.999927

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	02/13/2017	12:11
2	CAL2	0.05	1	50	50	0.032	02/13/2017	12:11
3	CAL3	0.10	1	50	50	0.062	02/13/2017	12:12
4	CAL4	0.30	1	50	50	0.196	02/13/2017	12:12
5	CAL5	0.50	1	50	50	0.332	02/13/2017	12:13
6	CAL6	1.00	1	50	50	0.651	02/13/2017	12:13



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB85801

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.490	02/13/2017	12:14
2	ICB		1	50	50	0.000	0.000	02/13/2017	12:14
3	CCV1	0.50	1	50	50	0.329	0.504	02/13/2017	12:15
4	CCB1		1	50	50	0.000	0.000	02/13/2017	12:15
5	PB96902BL		1	50	50	0.000	0.000	02/13/2017	12:16
6	PB96902BS	0.50	1	50	50	0.316	0.484	02/13/2017	12:16
7	I1773-01		1	50	50	0.046	0.071	02/13/2017	12:17
8	I1773-01DUP		1	50	50	0.047	0.072	02/13/2017	12:17
9	I1773-01MS	0.50	1	50	50	0.364	0.557	02/13/2017	12:18
10	I1773-01MSD	0.50	1	50	50	0.363	0.556	02/13/2017	12:18
11	CCV2	0.50	1	50	50	0.330	0.505	02/13/2017	12:19
12	CCB2		1	50	50	0.000	0.000	02/13/2017	12:19