



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB85808

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP52896
COD CCV std, 50ppm	WP52895

Temp In (C): <u>150</u>	Date In: <u>02/14/2017</u>	Time In: <u>09:35</u>
Temp Out (C): <u>150</u>	Date Out: <u>02/14/2017</u>	Time Out: <u>11:36</u>

Intercept: 0

Slope: 1

Regression: 1

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	02/06/2017	12:40
2	CAL2	10	1	Water	10.000	02/06/2017	12:40
3	CAL3	50	1	Water	50.000	02/06/2017	12:41
4	CAL4	100	1	Water	100.000	02/06/2017	12:41
5	CAL5	150	1	Water	150.000	02/06/2017	12:42



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB85808

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.000	02/06/2017	12:42
2	ICB				1	Water	0.000	0.000	02/06/2017	12:43
3	CCV1	50			1	Water	51.000	51.000	02/14/2017	12:43
4	CCB1				1	Water	0.000	0.000	02/14/2017	12:44
5	LB85808BL				1	Water	0.000	0.000	02/14/2017	12:44
6	LB85808BS	50			1	Water	52.000	52.000	02/14/2017	12:45
7	I1752-06				50	Water	42.000	42.000	02/14/2017	12:45
8	I1753-06				50	Water	51.000	51.000	02/14/2017	12:46
9	I1756-02				1	Water	2.000	2.000	02/14/2017	12:46
10	I1756-02DUP				1	Water	1.000	1.000	02/14/2017	12:47
11	I1756-02MS	50			1	Water	53.000	53.000	02/14/2017	12:47
12	I1756-02MSD	50			1	Water	52.000	52.000	02/14/2017	12:48
13	I1770-06				50	Water	55.000	55.000	02/14/2017	12:48
14	I1773-01				1	Water	2.000	2.000	02/14/2017	12:49
15	CCV2	50			1	Water	50.000	50.000	02/14/2017	12:49
16	CCB2				1	Water	0.000	0.000	02/14/2017	12:50
17	I1793-06				50	Water	44.000	44.000	02/14/2017	12:50
18	CCV3	50			1	Water	51.000	51.000	02/14/2017	12:51
19	CCB3				1	Water	0.000	0.000	02/14/2017	12:51

WORKLIST(Hardcopy Internal Chain)

LB85808

WorkList Name : COD020817

WorkList ID : 95460

Date : 2/8/2017 10:12:53 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I1752-06	COD	Conc H2SO4 1c	AMER08	G41	COD	02/07/2017	SM5220 D
	Water	I1753-06	COD	Conc H2SO4 1c	AMER08	G31	COD	02/06/2017	SM5220 D
	Water	I1770-06	COD	Conc H2SO4 1c	AMER08	H13	COD	02/08/2017	SM5220 D
	Water	I1793-06	COD	Conc H2SO4 1c	AMER08	H21	COD	02/09/2017	SM5220 D
	Water	I1773-01	COD	Conc H2SO4 1c	NAT104	H12	DSN001	02/08/2017	SM5220 D
	Water	I1756-02	COD	Conc H2SO4 1c	SHAW22	H11	OUTFALL001-170207	02/07/2017	SM5220 D

Date/Time 02/14/17 9:20
 Received by: IB
 Relinquished by: CS

Date/Time 02/14/17 10:30
 Received by: CS
 Relinquished by: IB