

LB85809

=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AL

2/14/2017 11:08

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.328	0.0	0.092	
ICB1	1.024	0.0	0.005	
CCV1	235.958	0.0	0.227	
CCB1	1.036	0.0	0.005	
PB96906BLW	1.090	0.0	0.005	
PB96906BSW	96.682	0.0	0.096	
LOWPB96906	10.836	0.0	0.015	
HIGHPB96906	487.563	0.0	0.465	
PB96907BLS	1.370	0.0	0.006	
PB96907BSS	96.125	0.0	0.095	
CCV2	242.022	0.0	0.233	
CCB2	1.003	0.0	0.005	
I1792-01	1.564	0.0	0.006	
I1792-01D	1.410	0.0	0.006	
I1792-01S	33.410	0.0	0.036	
I1792-01SD	34.204	0.0	0.037	
I1792-03	1.611	0.0	0.006	
I1790-01	1.551	0.0	0.006	
I1790-01D	1.022	0.0	0.005	
I1790-01S	33.257	0.0	0.036	
I1790-01SD	33.496	0.0	0.036	
CCV3	255.781	0.0	0.246	
CCB3	1.807	0.0	0.006	

N	23
Mean	72.441
SD	122.1880
CV%	168.67

LB 85809

Aquakem v. 7.2AQ1

Results from time period:

AK

Tue Feb 14 09:36:18 2017

Tue Feb 14 10:53:59 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.9853	µg/l	2/14/2017 9:36:18	
5.0PPBCN	A	Total CN	P	5.5592	µg/l	2/14/2017 9:36:19	
10PPBCN	A	Total CN	P	10.4879	µg/l	2/14/2017 9:36:20	
50PPBCN	A	Total CN	P	47.7403	µg/l	2/14/2017 9:36:21	
100PPBCN	A	Total CN	P	100.145	µg/l	2/14/2017 9:36:22	
250PPBCN	A	Total CN	P	249.8013	µg/l	2/14/2017 9:36:23	
500PPBCN	A	Total CN	P	500.2809	µg/l	2/14/2017 9:36:24	
ICV1	S	Total CN	P	92.3281	µg/l	2/14/2017 10:13:13	
ICB1	S	Total CN	P	1.0243	µg/l	2/14/2017 10:13:14	
CCV1	S	Total CN	P	235.9577	µg/l	2/14/2017 10:13:15	
CCB1	S	Total CN	P	1.0365	µg/l	2/14/2017 10:13:16	
PB96906BLW	S	Total CN	P	1.0896	µg/l	2/14/2017 10:13:17	
PB96906BSW	S	Total CN	P	96.6819	µg/l	2/14/2017 10:13:18	
LOWPB96906	S	Total CN	P	10.836	µg/l	2/14/2017 10:13:19	
HIGHPB96906	S	Total CN	P	487.5634	µg/l	2/14/2017 10:13:20	
PB96907BLS	S	Total CN	P	1.3699	µg/l	2/14/2017 10:20:47	
PB96907BSS	S	Total CN	P	96.125	µg/l	2/14/2017 10:20:48	
CCV2	S	Total CN	P	242.0219	µg/l	2/14/2017 10:20:49	
CCB2	S	Total CN	P	1.0026	µg/l	2/14/2017 10:20:50	
I1792-01	S	Total CN	P	1.5644	µg/l	2/14/2017 10:20:51	
I1792-01D	S	Total CN	P	1.4099	µg/l	2/14/2017 10:20:52	
I1792-01S	S	Total CN	P	33.4101	µg/l	2/14/2017 10:20:53	
I1792-01SD	S	Total CN	P	34.2042	µg/l	2/14/2017 10:20:54	
I1792-03	S	Total CN	P	1.611	µg/l	2/14/2017 10:20:55	
I1790-01	S	Total CN	P	1.5513	µg/l	2/14/2017 10:53:54	
I1790-01D	S	Total CN	P	1.0219	µg/l	2/14/2017 10:53:55	
I1790-01S	S	Total CN	P	33.2574	µg/l	2/14/2017 10:53:56	
I1790-01SD	S	Total CN	P	33.4959	µg/l	2/14/2017 10:53:57	
CCV3	S	Total CN	P	255.7815	µg/l	2/14/2017 10:53:58	
CCB3	S	Total CN	P	1.8074	µg/l	2/14/2017 10:53:59	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/14/2017 9:40


Test Total CN

Accepted 2/14/2017 9:40

Factor 1060
Bias 0.004

Coeff. of det. 0.999967

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	0.9853	0.0000	
2	5.0PPBCN	0.010	5.5592	5.0000	
3	10PPBCN	0.014	10.4879	10.0000	
4	50PPBCN	0.049	47.7403	50.0000	
5	100PPBCN	0.099	100.1450	100.0000	
6	250PPBCN	0.240	249.8013	250.0000	
7	500PPBCN	0.477	500.2809	500.0000	