

Chemtech Consulting Group


Analytical Review Report

Date Printed : 2/14/17

Analyst : AK

Data File : LB85809.csv

Approved By : 

Approved Date : 

Worksheet # :


m9012 A-B Total Amenable & Reactive cyanide -16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method				Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	2/14/17						
Cyanide		PASS	0.985	0.001							mg/L	
5.0PPBCN	5.0PPBCN				W	2/14/17						
Cyanide		PASS	5.559	0.006							mg/L	
10PPBCN	10PPBCN				W	2/14/17						
Cyanide		PASS	10.488	0.010							mg/L	
50PPBCN	50PPBCN				W	2/14/17						
Cyanide		PASS	47.740	0.048							mg/L	
100PPBCN	100PPBCN				W	2/14/17						
Cyanide		PASS	100.145	0.100							mg/L	
250PPBCN	250PPBCN				W	2/14/17						
Cyanide		PASS	249.801	0.250							mg/L	
500PPBCN	500PPBCN				W	2/14/17						
Cyanide		PASS	500.281	0.500							mg/L	
ICV1	ICV1				W	2/14/17						
Cyanide		PASS	92.328	0.09		90.0	90	110			mg/L	
ICB1	ICB1				W	2/14/17						
Cyanide		PASS	1.024	0.001				+/-0.0030			mg/L	
CCV1	CCV1				W	2/14/17						
Cyanide		PASS	235.958	0.24		96.0	90	110			mg/L	
CCB1	CCB1				W	2/14/17						
Cyanide		PASS	1.036	0.001				+/-0.0030			mg/L	
PB96906BLW	PB96906BLW				W	2/14/17						
Cyanide		PASS	1.090	0.001				+/-0.0030			mg/L	
PB96906BSW	PB96906BSW				W	2/14/17						
Cyanide		PASS	96.682	0.10		100.0	85	115			mg/L	
LOWPB96906	LOWPB96906				W	2/14/17						
Cyanide		PASS	10.836	0.011							mg/L	
HIGHPB96906	HIGHPB96906				W	2/14/17						
Cyanide		PASS	487.563	0.488							mg/L	
PB96907BLS	PB96907BLS				S	2/14/17						
Cyanide		PASS	1.370	0.069				+/-0.0330			mg/Kg	
PB96907BSS	PB96907BSS				S	2/14/17						
Cyanide		PASS	96.125	4.81		96.0	85.00	115.00			mg/Kg	
CCV2	CCV2				W	2/14/17						
Cyanide		PASS	242.022	0.24		96.0	90	110			mg/L	
CCB2	CCB2				W	2/14/17						
Cyanide		PASS	1.003	0.001				+/-0.0030			mg/L	
I1792-01	FK-GF-02-002-A				S	2/14/17						
Cyanide		PASS	1.564	0.081							mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/14/17
 Analyst : AK
 Data File : LB85809.csv

Approved By : 
 Approved Date :
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB	Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide											
I1792-01D	FK-GF-02-002-AD		1	S	2/14/17						
Cyanide	PASS	1.410	0.073					10.4	20	mg/Kg	
I1792-01S	FK-GF-02-002-AS		1	S	2/14/17						
Cyanide	PASS	33.410	1.750		80.0	75	125			mg/Kg	
I1792-01SD	FK-GF-02-002-ASD		1	S	2/14/17						
Cyanide	PASS	34.204	1.800		81.0	75	125	2.8	20	mg/Kg	
I1792-03	FK-GF-02-002-B		1	S	2/14/17						
Cyanide	PASS	1.611	0.084							mg/Kg	
I1790-01	38-FB021017		1	W	2/14/17						
Cyanide	PASS	1.551	0.002							mg/L	
I1790-01D	38-FB021017D		1	W	2/14/17						
Cyanide	PASS	1.022	0.001					0	20	mg/L	
I1790-01S	38-FB021017S		1	W	2/14/17						
Cyanide	PASS	33.257	0.033		83.0	75	125			mg/L	
I1790-01SD	38-FB021017SD		1	W	2/14/17						
Cyanide	PASS	33.496	0.034		85.0	75	125	3.0	20	mg/L	
CCV3	CCV3			W	2/14/17						
Cyanide	PASS	255.781	0.26		104.0	90	110			mg/L	
CCB3	CCB3			W	2/14/17						
Cyanide	PASS	1.807	0.002				+/-0.0030			mg/L	