


Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/14/17

Analyst : AK

Data File : LB85810.csv

Approved By : 

Approved Date : 2/14/17

Worksheet # :

msm4500 - a C.E - cyanide - 8

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	2/14/17						
Cyanide		PASS	0.985		0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	2/14/17						
Cyanide		PASS	5.559		0.006						mg/L	
10PPBCN	10PPBCN				W	2/14/17						
Cyanide		PASS	10.488		0.010						mg/L	
50PPBCN	50PPBCN				W	2/14/17						
Cyanide		PASS	47.740		0.048						mg/L	
100PPBCN	100PPBCN				W	2/14/17						
Cyanide		PASS	100.145		0.100						mg/L	
250PPBCN	250PPBCN				W	2/14/17						
Cyanide		PASS	249.801		0.250						mg/L	
500PPBCN	500PPBCN				W	2/14/17						
Cyanide		PASS	500.281		0.500						mg/L	
ICV1	ICV1				W	2/14/17						
Cyanide		PASS	93.459		0.09	90.0	85	115			mg/L	
ICB1	ICB1				W	2/14/17						
Cyanide		PASS	0.958		0.001			+/-0.0010			mg/L	
CCV1	CCV1				W	2/14/17						
Cyanide		PASS	240.269		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	2/14/17						
Cyanide		PASS	1.331		0.001			+/-0.0010			mg/L	
PB96905BLW	PB96905BLW				W	2/14/17						
Cyanide		PASS	1.403		0.001			+/-0.0010			mg/L	
PB96905BSW	PB96905BSW				W	2/14/17						
Cyanide		PASS	96.744		0.097	97.0	85	115			mg/L	
MIDPB96905	MIDPB96905				W	2/14/17						
Cyanide		PASS	243.813		0.244						mg/L	
I1789-01	EFF-WW		1		W	2/14/17						
Cyanide		PASS	1.709		0.002						mg/L	
I1789-01D	EFF-WWD		1		W	2/14/17						
Cyanide		FAIL	1.378		0.001				67	20	mg/L	
I1789-01S	EFF-WWS		1		W	2/14/17						
Cyanide		PASS	35.388		0.035	83.0	75	125			mg/L	
I1789-01SD	EFF-WWSD		1		W	2/14/17						
Cyanide		PASS	34.956		0.035	83.0	75	125	0	20	mg/L	
CCV2	CCV2				W	2/14/17						
Cyanide		PASS	255.658		0.26	104.0	90	110			mg/L	
CCB2	CCB2				W	2/14/17						
Cyanide		PASS	1.199		0.001			+/-0.0010			mg/L	