

7471A


LB 85812

INSTRUMENT ID: CV2

Linear

 μ Abs.:

46898



A= 0.0000e+000

B= 2.1524e-004

C= 2.5479e-002

Rho= 0.9997762

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.064	0.064	181	0.000	181				
0.2	0.200	0.217	0.017	891	0.0 %	891				
2.5	2.500	2.482	-0.018	11413	0.0 %	11413				
5.0	5.000	4.927	-0.073	22771	0.0 %	22771				
7.5	7.500	7.390	-0.110	34216	0.0 %	34216				
10.0	10.000	10.120	0.120	46898	0.0 %	46898				

LB85812 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 14 Feb 2017 14:05:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	181	-	0.0000	7471A	PPB	14 Feb 2017 14:11:40	S	Std
0.2 - 1	SD.2	891	-	0.2000	7471A	PPB	14 Feb 2017 14:13:58	S	Std
2.5 - 1	SD.5	11413	-	2.5000	7471A	PPB	14 Feb 2017 14:16:15	S	Std
5.0 - 1	SD	22771	-	5.0000	7471A	PPB	14 Feb 2017 14:18:33	S	Std
7.5 - 1	SD.5	34216	-	7.5000	7471A	PPB	14 Feb 2017 14:20:50	S	Std
10.0 - 1	SD	46898	-	10.0000	7471A	PPB	14 Feb 2017 14:30:58	S	Std
ICV31 - 1	ICV31	18022	3.9045	-7471A	PPB	14 Feb 2017 14:33:42	U	SMPL	
ICB31 - 1	ICB31	191	0.0666	-7471A	PPB	14 Feb 2017 14:35:57	U	SMPL	
CCV18 - 1	CCV18	22736	4.9191	-7471A	PPB	14 Feb 2017 14:38:14	U	SMPL	
CCB18 - 1	CCB18	194	0.0672	-7471A	PPB	14 Feb 2017 14:40:29	U	SMPL	
CRA - 1	CRA	906	0.2205	-7471A	PPB	14 Feb 2017 14:42:45	U	SMPL	
HighStd - 1	HighStd	47360	10.2191	-7471A	PPB	14 Feb 2017 14:45:01	U	SMPL	
ChkStd - 1	ChkStd	33273	7.1871	-7471A	PPB	14 Feb 2017 14:47:17	U	SMPL	
PB96908BL - 1	PBS	86	0.0440	-7471A	PPB	14 Feb 2017 14:49:33	U	SMPL	
PB96908BS - 1	LCSS	16147	3.5009	-7471A	PPB	14 Feb 2017 14:51:49	U	SMPL	
I1792-01 - 1	FK-GF-02-002-A	2717	0.6103	-7471A	PPB	14 Feb 2017 14:54:06	U	SMPL	
I1792-03 - 1	FK-GF-02-002-B	3203	0.7149	-7471A	PPB	14 Feb 2017 15:01:14	U	SMPL	
I1792-03DUP - 1	FK-GF-02-002-BDUP	3207	0.7157	-7471A	PPB	14 Feb 2017 15:03:31	U	SMPL	
I1792-03MS - 1	FK-GF-02-002-BMS	18844	4.0814	-7471A	PPB	14 Feb 2017 15:05:49	U	SMPL	
I1792-03MSD - 1	FK-GF-02-002-BMSD	19625	4.2495	-7471A	PPB	14 Feb 2017 15:08:05	U	SMPL	
CCV19 - 1	CCV19	22834	4.9402	-7471A	PPB	14 Feb 2017 15:10:20	U	SMPL	
CCB19 - 1	CCB19	-37	0.0175	-7471A	PPB	14 Feb 2017 15:12:36	U	SMPL	
I1794-04 - 1	22A-9-11	257	0.0808	-7471A	PPB	14 Feb 2017 15:14:52	U	SMPL	
I1816-01 - 1	H-1	998	0.2403	-7471A	PPB	14 Feb 2017 15:17:08	U	SMPL	
I1792-03LX5 - 1		625	0.1600	-7471A	PPB	14 Feb 2017 15:19:24	U	SMPL	
I1792-03A - 1		19651	4.2551	-7471A	PPB	14 Feb 2017 15:21:40	U	SMPL	
CCV20 - 1	CCV20	22894	4.9531	-7471A	PPB	14 Feb 2017 15:23:57	U	SMPL	
CCB20 - 1	CCB20	-25	0.0201	-7471A	PPB	14 Feb 2017 15:26:14	U	SMPL	