

7470A


LB85830

INSTRUMENT ID: CV2

Linear

 μ Abs.:

35719



A= 0.0000e+000

B= 2.7920e-004

C= 4.3785e-002

Rho= 0.9998875

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	8	0.000	8				
0.2	0.200	0.246	0.046	725	0.0 %	725				
2.5	2.500	2.391	-0.109	8406	0.0 %	8406				
5.0	5.000	4.969	-0.031	17642	0.0 %	17642				
7.5	7.500	7.531	0.031	26818	0.0 %	26818				
10.0	10.000	10.016	0.016	35719	0.0 %	35719				

LB85830 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Feb 2017 11:55:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	7470A	PPB	15 Feb 2017 12:02:03	S	Std
0.2 - 1	50.2	725	-	0.2000	7470A	PPB	15 Feb 2017 12:04:12	S	Std
2.5 - 1	52.5	8406	-	2.5000	7470A	PPB	15 Feb 2017 12:06:20	S	Std
5.0 - 1	55	17642	-	5.0000	7470A	PPB	15 Feb 2017 12:08:28	S	Std
7.5 - 1	57.5	26818	-	7.5000	7470A	PPB	15 Feb 2017 12:15:36	S	Std
10.0 - 1	50	35719	-	10.0000	7470A	PPB	15 Feb 2017 12:22:50	S	Std
ICV32 - 1	ICV32	13481	3.8076	-7470A	PPB	PPB	15 Feb 2017 12:28:20	U	SMPL
ICB32 - 1	ICB32	-10	0.0410	-7470A	PPB	PPB	15 Feb 2017 12:30:26	U	SMPL
CCV21 - 1	CCV21	17449	4.9155	-7470A	PPB	PPB	15 Feb 2017 12:32:32	U	SMPL
CCB21 - 1	CCB21	10	0.0466	-7470A	PPB	PPB	15 Feb 2017 12:34:38	U	SMPL
CRA - 1	CRA	672	0.2314	-7470A	PPB	PPB	15 Feb 2017 12:36:45	U	SMPL
HighStd - 1	HighStd	36010	10.0977	-7470A	PPB	PPB	15 Feb 2017 12:38:52	U	SMPL
ChkStd - 1	ChkStd	25591	7.1887	-7470A	PPB	PPB	15 Feb 2017 12:40:59	U	SMPL
PB96932BL - 1	PBW	-58	0.0276	-7470A	PPB	PPB	15 Feb 2017 12:43:06	U	SMPL
PB96932BS - 1	LCSW	13583	3.8361	-7470A	PPB	PPB	15 Feb 2017 12:45:14	U	SMPL
PB96932TB - 1	PB96903TB	-28	0.0360	-7470A	PPB	PPB	15 Feb 2017 12:47:21	U	SMPL
I1790-02 - 1	38-WCS021017-0-15	2	0.0443	-7470A	PPB	PPB	15 Feb 2017 12:49:30	U	SMPL
I1790-02DUP - 1	38-WCS021017-0-15DUP	4	0.0449	-7470A	PPB	PPB	15 Feb 2017 12:51:38	U	SMPL
I1790-02MS - 1	38-WCS021017-0-15MS	13588	3.8375	-7470A	PPB	PPB	15 Feb 2017 12:53:46	U	SMPL
I1790-02MSD - 1	38-WCS021017-0-15MSD	13104	3.7024	-7470A	PPB	PPB	15 Feb 2017 12:55:53	U	SMPL
CCV22 - 1	CCV22	17573	4.9501	-7470A	PPB	PPB	15 Feb 2017 12:58:00	U	SMPL
CCB22 - 1	CCB22	-10	0.0410	-7470A	PPB	PPB	15 Feb 2017 13:00:07	U	SMPL
PB96933BL - 1	PBW	-9	0.0413	-7470A	PPB	PPB	15 Feb 2017 13:02:14	U	SMPL
PB96933BS - 1	LCSW	14985	4.2275	-7470A	PPB	PPB	15 Feb 2017 13:04:21	U	SMPL
PB96933TB - 1	PB96910TB	-15	0.0396	-7470A	PPB	PPB	15 Feb 2017 13:06:28	U	SMPL
I1794-06 - 1	22A-9-11	79	0.0658	-7470A	PPB	PPB	15 Feb 2017 13:08:35	U	SMPL
I1794-06DUP - 1	22A-9-11DUP	85	0.0675	-7470A	PPB	PPB	15 Feb 2017 13:10:42	U	SMPL
I1794-06MS - 1	22A-9-11MS	14027	3.9601	-7470A	PPB	PPB	15 Feb 2017 13:12:50	U	SMPL
I1794-06MSD - 1	22A-9-11MSD	14357	4.0522	-7470A	PPB	PPB	15 Feb 2017 13:14:58	U	SMPL
CCV23 - 1	CCV23	17487	4.9261	-7470A	PPB	PPB	15 Feb 2017 13:17:06	U	SMPL
CCB23 - 1	CCB23	-13	0.0402	-7470A	PPB	PPB	15 Feb 2017 13:19:14	U	SMPL
I1790-02LX5 - 1		-9	0.0413	-7470A	PPB	PPB	15 Feb 2017 13:21:21	U	SMPL
I1790-02A - 1		13742	3.8805	-7470A	PPB	PPB	15 Feb 2017 13:23:28	U	SMPL
I1794-06LX5 - 1		-1	0.0435	-7470A	PPB	PPB	15 Feb 2017 13:25:35	U	SMPL
I1794-06A - 1		14458	4.0804	-7470A	PPB	PPB	15 Feb 2017 13:27:42	U	SMPL
CCV24 - 1	CCV24	17372	4.8940	-7470A	PPB	PPB	15 Feb 2017 13:29:49	U	SMPL
CCB24 - 1	CCB24	-33	0.0346	-7470A	PPB	PPB	15 Feb 2017 13:31:55	U	SMPL