



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB85832

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
PH 10.01 BUFFER,COLOR CD 475ML	W2154
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/15/2017	14:30
2	CAL2	1	Water			7.01	02/15/2017	14:31
3	CAL3	1	Water			10.02	02/15/2017	14:32
4	ICV	1	Water			7.02	02/15/2017	14:33
5	CCV1	1	Water			2.01	02/15/2017	14:34
6	I1850-06	1	Solid	20.02	20	8.89	02/15/2017	14:35
7	I1859-05	1	Solid	20.04	20	9.14	02/15/2017	14:36
8	I1859-10	1	Solid	20.03	20	9.18	02/15/2017	14:37
9	I1859-15	1	Solid	20.03	20	8.76	02/15/2017	14:38
10	I1859-15DUP	1	Solid	20.05	20	8.77	02/15/2017	14:39
11	CCV2	1	Water			12.01	02/15/2017	14:40

1385832

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1850 WorkList ID : 95642 Date : 2/15/2017 1:15:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I1850-06	Corrosivity	Cool 4 deg C	DEWB01	H31	TP21-COMP15	02/15/2017	9045C
02/20/2017	Solid	I1859-05	Corrosivity	Cool 4 deg C	PSEG01	F42	280514-COMP-0-2FT	02/15/2017	9045C
02/20/2017	Solid	I1859-10	Corrosivity	Cool 4 deg C	PSEG01	F42	280517-COMP-0-2FT	02/15/2017	9045C
02/20/2017	Solid	I1859-15	Corrosivity	Cool 4 deg C	PSEG01	F42	280385-COMP-0-2FT	02/15/2017	9045C

Date/Time 02-15-17 1:15 PM
Received by: JH
Relinquished by:

Date/Time 02-15-17 5:30 PM
Received by: [Signature]
Relinquished by: JP