


LB85836

7471A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.1868e-004

C= -1.9544e-002

Rho= 0.9999160

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	26	0.000	26				
0.2	0.200	0.184	-0.016	931	0.0 %	931				
2.5	2.500	2.505	0.005	11544	0.0 %	11544				
5.0	5.000	5.090	0.090	23366	0.0 %	23366				
7.5	7.500	7.429	-0.071	34060	0.0 %	34060				
10.0	10.000	10.006	0.006	45847	0.0 %	45847				

LB85836 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 Feb 2017 14:47:02

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	26	-	0.00007471A	PPB	15 Feb 2017 14:48:55	S	Std
0.2 - 1	50.2	931	-	0.20007471A	PPB	15 Feb 2017 14:51:13	S	Std
2.5 - 1	52.5	11544	-	2.50007471A	PPB	15 Feb 2017 14:53:30	S	Std
5.0 - 1	55	23366	-	5.00007471A	PPB	15 Feb 2017 14:55:48	S	Std
7.5 - 1	57.5	34060	-	7.50007471A	PPB	15 Feb 2017 14:58:05	S	Std
10.0 - 1	3/0	45847	-	10.00007471A	PPB	15 Feb 2017 15:02:56	S	Std
ICV33 - 1	ICV33	16867	3.6689	-7471A	PPB	15 Feb 2017 15:05:34	U	SMPL
ICB33 - 1	ICB33	-13	-0.0224	-7471A	PPB	15 Feb 2017 15:07:49	U	SMPL
CCV25 - 1	CCV25	22875	4.9827	-7471A	PPB	15 Feb 2017 15:10:05	U	SMPL
CCB25 - 1	CCB25	-26	-0.0252	-7471A	PPB	15 Feb 2017 15:12:20	U	SMPL
CRA - 1	CRA	927	0.1832	-7471A	PPB	15 Feb 2017 15:14:36	U	SMPL
HighStd - 1	HighStd	45837	10.0040	-7471A	PPB	15 Feb 2017 15:16:52	U	SMPL
ChkStd - 1	ChkStd	31871	6.9500	-7471A	PPB	15 Feb 2017 15:19:08	U	SMPL
PB96948BL - 1	PBS	-57	-0.0320	-7471A	PPB	15 Feb 2017 15:21:24	U	SMPL
PB96948BS - 1	LCSS	17210	3.7439	-7471A	PPB	15 Feb 2017 15:28:50	U	SMPL
I1822-02 - 1	38-BC-01-0-2	9894	2.1441	-7471A	PPB	15 Feb 2017 15:31:06	U	SMPL
I1822-03 - 1	38-SB-01-65-70	56	-0.0073	-7471A	PPB	15 Feb 2017 15:33:23	U	SMPL
I1825-01 - 1	AIRSIDELIFTSAMPLE1C	1819	0.3782	-7471A	PPB	15 Feb 2017 15:35:41	U	SMPL
I1826-03 - 1	TP1-2-COMP14	22367	4.8716	-7471A	PPB	15 Feb 2017 15:37:59	U	SMPL
I1835-01 - 1	TEST-PIT-3-COMP	1180	0.2385	-7471A	PPB	15 Feb 2017 15:40:14	U	SMPL
CCV26 - 1	CCV26	23140	5.0407	-7471A	PPB	15 Feb 2017 15:42:30	U	SMPL
CCB26 - 1	CCB26	-6	-0.0209	-7471A	PPB	15 Feb 2017 15:44:46	U	SMPL
I1836-01 - 1	TEST-PIT-A-COMP	618	0.1156	-7471A	PPB	15 Feb 2017 15:47:02	U	SMPL
I1837-01 - 1	P-1	158	0.0150	-7471A	PPB	15 Feb 2017 15:49:18	U	SMPL
I1838-01 - 1	FK-BPM-01-BB9-BB10-BB11-BB12	421	0.0725	-7471A	PPB	15 Feb 2017 15:51:34	U	SMPL
I1839-01 - 1	FK-BPM-01-AA9B-AA10B-AA11B-AA12B	407	0.0695	-7471A	PPB	15 Feb 2017 15:53:50	U	SMPL
I1839-01DUP - 1	FK-BPM-01-AA9B-AA10B-AA11B-AA12BDUP	378	0.0631	-7471A	PPB	15 Feb 2017 16:01:00	U	SMPL
I1839-01MS - 1	FK-BPM-01-AA9B-AA10B-AA11B-AA12BMS	16823	3.6593	-7471A	PPB	15 Feb 2017 16:03:17	U	SMPL
I1839-01MSD - 1	FK-BPM-01-AA9B-AA10B-AA11B-AA12BMSD	17089	3.7174	-7471A	PPB	15 Feb 2017 16:05:34	U	SMPL
I1839-01LX5 - 1		70	-0.0042	-7471A	PPB	15 Feb 2017 16:07:52	U	SMPL
I1839-01A - 1		17351	3.7747	-7471A	PPB	15 Feb 2017 16:10:10	U	SMPL
CCV27 - 1	CCV27	23055	5.0221	-7471A	PPB	15 Feb 2017 16:12:26	U	SMPL
CCB27 - 1	CCB27	-66	-0.0340	-7471A	PPB	15 Feb 2017 16:14:42	U	SMPL