


7470A

LB85849
INSTRUMENT ID: CV4

Linear

 μ Abs.:

37627



A= 0.0000e+000

B= 2.6479e-004

C= -1.8524e-002

Rho= 0.9999294

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	9	0.000	9				
0.2	0.200	0.180	-0.020	750	0.0 %	750				
2.5	2.500	2.486	-0.014	9457	0.0 %	9457				
5.0	5.000	5.085	0.085	19273	0.0 %	19273				
7.5	7.500	7.521	0.021	28473	0.0 %	28473				
10.0	10.000	9.945	-0.055	37627	0.0 %	37627				

LB85849 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Feb 2017 11:59:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000	7470A	PPB	16 Feb 2017 12:08:11	S	Std
0.2 - 1	50.2	750	-	0.2000	7470A	PPB	16 Feb 2017 12:10:20	S	Std
2.5 - 1	52.5	9457	-	2.5000	7470A	PPB	16 Feb 2017 12:12:28	S	Std
5.0 - 1	55	19273	-	5.0000	7470A	PPB	16 Feb 2017 12:14:37	S	Std
7.5 - 1	57.5	28473	-	7.5000	7470A	PPB	16 Feb 2017 12:16:44	S	Std
10.0 - 1	510	37627	-	10.0000	7470A	PPB	16 Feb 2017 12:18:51	S	Std
ICV34 - 1	ICV34	14541	3.8318	-	7470A	PPB	16 Feb 2017 12:21:27	U	SMPL
ICB34 - 1	ICB34	-39	-0.0289	-	7470A	PPB	16 Feb 2017 12:23:33	U	SMPL
CCV28 - 1	CCV28	18730	4.9410	-	7470A	PPB	16 Feb 2017 12:25:40	U	SMPL
CCB28 - 1	CCB28	-30	-0.0265	-	7470A	PPB	16 Feb 2017 12:27:46	U	SMPL
CRA - 1	CRA	726	0.1737	-	7470A	PPB	16 Feb 2017 12:29:53	U	SMPL
HighStd - 1	HighStd	37284	9.8540	-	7470A	PPB	16 Feb 2017 12:31:59	U	SMPL
ChkStd - 1	ChkStd	26330	6.9534	-	7470A	PPB	16 Feb 2017 12:34:06	U	SMPL
PB96971BL - 1	PBW	-75	-0.0384	-	7470A	PPB	16 Feb 2017 12:36:13	U	SMPL
PB96971BS - 1	LCSW	14290	3.7653	-	7470A	PPB	16 Feb 2017 12:38:21	U	SMPL
PB96971TB - 1	PB96935TB	-40	-0.0291	-	7470A	PPB	16 Feb 2017 12:40:29	U	SMPL
I1822-01 - 1	38-WCS021317-15-70	-14	-0.0222	-	7470A	PPB	16 Feb 2017 12:42:37	U	SMPL
I1825-02 - 1	AIRSIDELIFTSAMPLE1C	-5	-0.0198	-	7470A	PPB	16 Feb 2017 12:44:45	U	SMPL
I1826-04 - 1	TP1-2-COMP14	18	-0.0138	-	7470A	PPB	16 Feb 2017 12:46:54	U	SMPL
I1832-02 - 1	WC-B1-002	219	0.0395	-	7470A	PPB	16 Feb 2017 12:49:00	U	SMPL
CCV29 - 1	CCV29	18459	4.8693	-	7470A	PPB	16 Feb 2017 12:51:07	U	SMPL
CCB29 - 1	CCB29	-17	-0.0230	-	7470A	PPB	16 Feb 2017 12:53:14	U	SMPL
I1835-05 - 1	TEST-PIT-3-COMP	-11	-0.0214	-	7470A	PPB	16 Feb 2017 12:55:21	U	SMPL
I1836-04 - 1	TEST-PIT-A-COMP	-3	-0.0193	-	7470A	PPB	16 Feb 2017 12:57:28	U	SMPL
I1836-04DUP - 1	TEST-PIT-A-COMPDUP	3	-0.0177	-	7470A	PPB	16 Feb 2017 12:59:35	U	SMPL
I1836-04MS - 1	TEST-PIT-A-COMPMS	13404	3.5307	-	7470A	PPB	16 Feb 2017 13:01:42	U	SMPL
I1836-04MSD - 1	TEST-PIT-A-COMPMSD	13413	3.5331	-	7470A	PPB	16 Feb 2017 13:03:50	U	SMPL
PB96974BL - 1	PBW	-26	-0.0254	-	7470A	PPB	16 Feb 2017 13:05:57	U	SMPL
PB96974BS - 1	LCSW	13841	3.6465	-	7470A	PPB	16 Feb 2017 13:08:06	U	SMPL
I1790-01 - 1	38-FB021017	-47	-0.0310	-	7470A	PPB	16 Feb 2017 13:10:14	U	SMPL
I1856-01 - 1	SAMPLE-1	-21	-0.0241	-	7470A	PPB	16 Feb 2017 13:12:22	U	SMPL
I1856-01DUP - 1	SAMPLE-1DUP	-12	-0.0217	-	7470A	PPB	16 Feb 2017 13:14:29	U	SMPL
CCV30 - 1	CCV30	18746	4.9453	-	7470A	PPB	16 Feb 2017 13:16:36	U	SMPL
CCB30 - 1	CCB30	-26	-0.0254	-	7470A	PPB	16 Feb 2017 13:18:43	U	SMPL
I1856-01MS - 1	SAMPLE-1MS	14449	3.8074	-	7470A	PPB	16 Feb 2017 13:20:50	U	SMPL
I1856-01MSD - 1	SAMPLE-1MSD	13809	3.6380	-	7470A	PPB	16 Feb 2017 13:22:57	U	SMPL
I1856-02 - 1	SAMPLE-2	-49	-0.0315	-	7470A	PPB	16 Feb 2017 13:25:04	U	SMPL
I1836-04LX5 - 1		-11	-0.0214	-	7470A	PPB	16 Feb 2017 13:27:10	U	SMPL
I1836-04A - 1		14611	3.8503	-	7470A	PPB	16 Feb 2017 13:29:18	U	SMPL
I1856-01LX5 - 1		-31	-0.0267	-	7470A	PPB	16 Feb 2017 13:31:26	U	SMPL
I1856-01A - 1		14699	3.8736	-	7470A	PPB	16 Feb 2017 13:33:34	U	SMPL
CCV31 - 1	CCV31	18203	4.8015	-	7470A	PPB	16 Feb 2017 13:35:42	U	SMPL
CCB31 - 1	CCB31	-48	-0.0312	-	7470A	PPB	16 Feb 2017 13:37:50	U	SMPL