



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85855

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/16/2017	14:10
2	CAL2	1	Water			7.00	02/16/2017	14:11
3	CAL3	1	Water			10.01	02/16/2017	14:12
4	ICV	1	Water			6.99	02/16/2017	14:13
5	CCV1	1	Water			2.02	02/16/2017	14:15
6	I1865-02	1	Solid	20.03	20	7.66	02/16/2017	14:16
7	I1865-02DUP	1	Solid	20.04	20	7.67	02/16/2017	14:17
8	I1869-02	1	Solid	20.05	20	9.31	02/16/2017	14:18
9	CCV2	1	Water			12.02	02/16/2017	14:19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I1865 WorkList ID : 95685 Date : 2/16/2017 1:42:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/23/2017	Solid	I1865-02	Corrosivity	Cool 4 deg C	OSCI01		TILCON-SCREENINGS	02/15/2017	9045C
02/23/2017	Solid	I1869-02	Corrosivity	Cool 4 deg C	OSCI01	H22	TILCON-NO-57	02/15/2017	9045C

Date/Time 02-16-17 1:42 PM
 Received by: TV
 Relinquished by: QR

Date/Time 02-16-17 5:30 PM
 Received by: CP
 Relinquished by: JP