



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB85868

Temp.: 22.2 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	02/17/2017	09:30
2	CAL2	1	Water			7.01	02/17/2017	09:31
3	CAL3	1	Water			10.01	02/17/2017	09:32
4	ICV	1	Water			6.99	02/17/2017	09:33
5	CCV1	1	Water			2.02	02/17/2017	09:34
6	I1884-13	1	Water			6.62	02/17/2017	09:35
7	I1884-16	1	Water			6.51	02/17/2017	09:36
8	I1884-19	1	Water			6.38	02/17/2017	09:37
9	I1884-19DUP	1	Water			6.39	02/17/2017	09:38
10	CCV2	1	Water			12.02	02/17/2017	09:39

1885868

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH11884

WorkList ID : 95713

Date : 2/17/2017 9:04:10 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/16/2017	Water	I1884-13-D	pH	Cool 4 deg C	PSEG01	F51	MANHOLE-001	02/16/2017	9045C
02/16/2017	Water	I1884-16-D	pH	Cool 4 deg C	PSEG01	F51	MANHOLE-002	02/16/2017	9045C
02/16/2017	Water	I1884-19-D	pH	Cool 4 deg C	PSEG01	F51	GROUNDWATER	02/16/2017	9045C

Date/Time 02-17-17 11:10 AM
Received by: 711
Relinquished by: [Signature]

Date/Time 02-17-17 11:10 AM
Received by: CP
Relinquished by: SP