


7470A

LIB 85873
INSTRUMENT ID: CVD

Linear

 μ Abs.:

37856



A= 0.0000e+000

B= 2.6400e-004

C= 1.6691e-002

Rho= 0.9999591

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	3	0.000	3				
0.2	0.200	0.210	0.010	731	0.0 %	731				
2.5	2.500	2.502	0.002	9415	0.0 %	9415				
5.0	5.000	4.928	-0.072	18604	0.0 %	18604				
7.5	7.500	7.532	0.032	28466	0.0 %	28466				
10.0	10.000	10.011	0.011	37856	0.0 %	37856				

LB85873 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Feb 2017 10:42:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	7470A	PPB	17 Feb 2017 10:47:01	S	Std
0.2 - 1	50.2	731	-	0.2000	7470A	PPB	17 Feb 2017 10:49:09	S	Std
2.5 - 1	52.5	9415	-	2.5000	7470A	PPB	17 Feb 2017 10:51:17	S	Std
5.0 - 1	55	18604	-	5.0000	7470A	PPB	17 Feb 2017 10:53:25	S	Std
7.5 - 1	57.5	28466	-	7.5000	7470A	PPB	17 Feb 2017 10:55:33	S	Std
10.0 - 1	510	37856	-	10.0000	7470A	PPB	17 Feb 2017 10:57:40	S	Std
ICV35 - 1	ICV35	14844	3.9355	-	7470A	PPB	17 Feb 2017 11:00:18	U	SMPL
ICB35 - 1	ICB35	-53	0.0027	-	7470A	PPB	17 Feb 2017 11:02:25	U	SMPL
CCV32 - 1	CCV32	18774	4.9730	-	7470A	PPB	17 Feb 2017 11:04:31	U	SMPL
CCB32 - 1	CCB32	-48	0.0040	-	7470A	PPB	17 Feb 2017 11:06:38	U	SMPL
CRA - 1	CRA	739	0.2118	-	7470A	PPB	17 Feb 2017 11:08:44	U	SMPL
HighStd - 1	HighStd	37639	9.9534	-	7470A	PPB	17 Feb 2017 11:10:50	U	SMPL
ChkStd - 1	ChkStd	26285	6.9559	-	7470A	PPB	17 Feb 2017 11:12:57	U	SMPL
PB96987BL - 1	PBW	-59	0.0011	-	7470A	PPB	17 Feb 2017 11:15:05	U	SMPL
PB96987BS - 1	LCSW	13857	3.6749	-	7470A	PPB	17 Feb 2017 11:17:12	U	SMPL
PB96987TB - 1	PB96972TB	-15	0.0127	-	7470A	PPB	17 Feb 2017 11:19:20	U	SMPL
I1850-06 - 1	TP21-COMP15	24	0.0230	-	7470A	PPB	17 Feb 2017 11:21:28	U	SMPL
I1850-06DUP - 1	TP21-COMP15DUP	28	0.0241	-	7470A	PPB	17 Feb 2017 11:23:36	U	SMPL
I1850-06MS - 1	TP21-COMP15MS	14124	3.7454	-	7470A	PPB	17 Feb 2017 11:25:45	U	SMPL
I1850-06MSD - 1	TP21-COMP15MSD	13883	3.6818	-	7470A	PPB	17 Feb 2017 11:27:52	U	SMPL
CCV33 - 1	CCV33	18603	4.9279	-	7470A	PPB	17 Feb 2017 11:29:59	U	SMPL
CCB33 - 1	CCB33	-49	0.0038	-	7470A	PPB	17 Feb 2017 11:32:06	U	SMPL
I1859-05 - 1	280514-COMP-0-2FT	-11	0.0138	-	7470A	PPB	17 Feb 2017 11:34:13	U	SMPL
I1859-10 - 1	280517-COMP-0-2FT	-20	0.0114	-	7470A	PPB	17 Feb 2017 11:36:20	U	SMPL
I1859-15 - 1	280385-COMP-0-2FT	-20	0.0114	-	7470A	PPB	17 Feb 2017 11:38:26	U	SMPL
I1850-06LX5 - 1		-6	0.0151	-	7470A	PPB	17 Feb 2017 11:40:33	U	SMPL
I1850-06A - 1		14753	3.9115	-	7470A	PPB	17 Feb 2017 11:42:41	U	SMPL
CCV34 - 1	CCV34	18771	4.9722	-	7470A	PPB	17 Feb 2017 11:44:48	U	SMPL
CCB34 - 1	CCB34	-33	0.0080	-	7470A	PPB	17 Feb 2017 11:46:57	U	SMPL