



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85875

Temp.: 21.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/17/2017	11:50
2	CAL2	1	Water			7.02	02/17/2017	11:51
3	CAL3	1	Water			10.02	02/17/2017	11:52
4	ICV	1	Water			7.01	02/17/2017	11:53
5	CCV1	1	Water			2.01	02/17/2017	11:54
6	I1884-14	1	Water			6.72	02/17/2017	11:55
7	I1884-14DUP	1	Water			6.73	02/17/2017	11:56
8	I1884-17	1	Water			6.45	02/17/2017	11:57
9	I1884-20	1	Water			6.16	02/17/2017	11:58
10	CCV2	1	Water			12.02	02/17/2017	12:00

WORKLIST(Hardcopy Internal Chain)

1085875

WorkList Name : CORROSIVITYW11884 WorkList ID : 95730 Date : 2/17/2017 11:37:35 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/16/2017	Water	I1884-14-1	Corrosivity		PSEG01		MANHOLE-001	02/16/2017	9040C
02/16/2017	Water	I1884-17-1	Corrosivity		PSEG01		MANHOLE-002	02/16/2017	9040C
02/16/2017	Water	I1884-20-1	Corrosivity		PSEG01		GROUNDWATER	02/16/2017	9040C

Date/Time 02-17-17 11:40 AM
 Received by: JH
 Relinquished by: CS

Date/Time 02-17-17 1:30 PM
 Received by: JP
 Relinquished by: JS