


7471A

LB85876
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47783



A= 0.0000e+000

B= 2.0930e-004

C= -3.0245e-002

Rho= 0.9999602

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	0	0.000	0				
0.2	0.200	0.173	-0.027	973	0.0 %	973				
2.5	2.500	2.556	0.056	12357	0.0 %	12357				
5.0	5.000	5.030	0.030	24175	0.0 %	24175				
7.5	7.500	7.500	0.000	35980	0.0 %	35980				
10.0	10.000	9.971	-0.029	47783	0.0 %	47783				

LB85876 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 17 Feb 2017 12:10:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	0	-	0.0000	7471A	PPB	17 Feb 2017 12:13:08	S	Std
0.2 - 1	50.2	973	-	0.2000	7471A	PPB	17 Feb 2017 12:15:26	S	Std
2.5 - 1	52.5	12357	-	2.5000	7471A	PPB	17 Feb 2017 12:17:44	S	Std
5.0 - 1	55	24175	-	5.0000	7471A	PPB	17 Feb 2017 12:20:01	S	Std
7.5 - 1	59.5	35980	-	7.5000	7471A	PPB	17 Feb 2017 12:22:18	S	Std
10.0 - 1	510	47783	-	10.0000	7471A	PPB	17 Feb 2017 12:24:35	S	Std
ICV36 - 1	ICV36	18720	3.8879	-7471A	PPB	17 Feb 2017 12:28:00	U	SMPL	
ICB36 - 1	ICB36	-22	-0.0349	-7471A	PPB	17 Feb 2017 12:30:15	U	SMPL	
CCV35 - 1	CCV35	24002	4.9934	-7471A	PPB	17 Feb 2017 12:35:05	U	SMPL	
CCB35 - 1	CCB35	-20	-0.0344	-7471A	PPB	17 Feb 2017 12:37:20	U	SMPL	
CRA - 1	CRA	975	0.1738	-7471A	PPB	17 Feb 2017 12:39:36	U	SMPL	
HighStd - 1	HighStd	48090	10.0350	-7471A	PPB	17 Feb 2017 12:41:52	U	SMPL	
ChkStd - 1	ChkStd	33368	6.9537	-7471A	PPB	17 Feb 2017 12:44:07	U	SMPL	
PB96983BL - 1	PBS	-65	-0.0438	-7471A	PPB	17 Feb 2017 12:46:24	U	SMPL	
PB96983BS - 1	LCSS	16650	3.4546	-7471A	PPB	17 Feb 2017 12:48:41	U	SMPL	
I1850-05 - 1	TP21-COMP15	26076	5.4275	-7471A	PPB	17 Feb 2017 12:50:57	U	SMPL	
I1859-01 - 1	280514-COMP-0-2FT	9519	1.9621	-7471A	PPB	17 Feb 2017 12:53:15	U	SMPL	
I1859-06 - 1	280517-COMP-0-2FT	4234	0.8559	-7471A	PPB	17 Feb 2017 12:55:32	U	SMPL	
I1859-11 - 1	280385-COMP-0-2FT	1321	0.2462	-7471A	PPB	17 Feb 2017 12:57:50	U	SMPL	
I1859-11DUP - 1	280385-COMP-0-2FTDUP	1403	0.2634	-7471A	PPB	17 Feb 2017 13:02:46	U	SMPL	
CCV36 - 1	CCV36	23033	4.7906	-7471A	PPB	17 Feb 2017 13:05:02	U	SMPL	
CCB36 - 1	CCB36	-19	-0.0342	-7471A	PPB	17 Feb 2017 13:07:18	U	SMPL	
I1859-11MS - 1	280385-COMP-0-2FTMS	17750	3.6848	-7471A	PPB	17 Feb 2017 13:09:34	U	SMPL	
I1859-11MSD - 1	280385-COMP-0-2FTMSD	17301	3.5909	-7471A	PPB	17 Feb 2017 13:11:50	U	SMPL	
I1865-01 - 1	TILCON-SCREENINGS	51	-0.0196	-7471A	PPB	17 Feb 2017 13:14:06	U	SMPL	
I1868-01 - 1	DRUM-AB-COMP	3017	0.6012	-7471A	PPB	17 Feb 2017 13:16:22	U	SMPL	
I1869-01 - 1	TILCON-NO-57	29	-0.0242	-7471A	PPB	17 Feb 2017 13:18:39	U	SMPL	
I1875-01 - 1	POSTEX-30-20170215	7850	1.6128	-7471A	PPB	17 Feb 2017 13:20:56	U	SMPL	
I1878-03 - 1	B1-2	115177	24.0764	-7471A	PPB	17 Feb 2017 13:23:14	U	SMPL	
I1878-06 - 1	B1-5	180709	37.7923	-7471A	PPB	17 Feb 2017 13:25:31	U	SMPL	
I1879-01 - 1	M-1	2975	0.5924	-7471A	PPB	17 Feb 2017 13:27:50	U	SMPL	
I1879-03 - 1	M-2	1411	0.2651	-7471A	PPB	17 Feb 2017 13:30:08	U	SMPL	
CCV37 - 1	CCV37	24087	5.0112	-7471A	PPB	17 Feb 2017 13:32:24	U	SMPL	
CCB37 - 1	CCB37	-26	-0.0357	-7471A	PPB	17 Feb 2017 13:34:41	U	SMPL	
I1879-05 - 1	M-3	1343	0.2508	-7471A	PPB	17 Feb 2017 13:36:57	U	SMPL	
I1879-07 - 1	M-4	642	0.1041	-7471A	PPB	17 Feb 2017 13:39:13	U	SMPL	
I1859-11LX5 - 1		279	0.0281	-7471A	PPB	17 Feb 2017 13:41:30	U	SMPL	
I1859-11A - 1		20223	4.2024	-7471A	PPB	17 Feb 2017 13:43:46	U	SMPL	
CCV38 - 1	CCV38	24015	4.9961	-7471A	PPB	17 Feb 2017 13:46:03	U	SMPL	
CCB38 - 1	CCB38	-30	-0.0365	-7471A	PPB	17 Feb 2017 13:48:20	U	SMPL	
I1878-03DLX5 - 1		22965	4.7763	-7471A	PPB	17 Feb 2017 13:54:06	U	SMPL	
I1878-06DLX5 - 1		37519	7.8225	-7471A	PPB	17 Feb 2017 13:56:23	U	SMPL	
CCV39 - 1	CCV39	24126	5.0193	-7471A	PPB	17 Feb 2017 13:58:41	U	SMPL	
CCB39 - 1	CCB39	-38	-0.0382	-7471A	PPB	17 Feb 2017 14:00:57	U	SMPL	