



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB85878

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/17/2017	13:10
2	CAL2	1	Water			7.02	02/17/2017	13:11
3	CAL3	1	Water			10.04	02/17/2017	13:12
4	ICV	1	Water			7.01	02/17/2017	13:13
5	CCV1	1	Water			2.02	02/17/2017	13:15
6	I1884-03	1	Solid	20.04	20	8.39	02/17/2017	13:16
7	I1884-03DUP	1	Solid	20.02	20	8.40	02/17/2017	13:17
8	I1884-06	1	Solid	20.05	20	9.06	02/17/2017	13:18
9	I1884-09	1	Solid	20.01	20	7.96	02/17/2017	13:20
10	I1884-12	1	Solid	20.04	20	8.48	02/17/2017	13:21
11	I1892-06	1	Solid	20.04	20	8.74	02/17/2017	13:22
12	CCV2	1	Water			12.01	02/17/2017	13:23