


7470A

LB85912
INSTRUMENT ID: CV4

Linear

 μ Abs.:

37276



A= 0.0000e+000

B= 2.6859e-004

C= -1.7827e-002

Rho= 0.9999612

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	-10	0.000	-10				
0.2	0.200	0.172	-0.028	707	0.0 %	707				
2.5	2.500	2.570	0.070	9634	0.0 %	9634				
5.0	5.000	5.000	0.000	18683	0.0 %	18683				
7.5	7.500	7.484	-0.016	27932	0.0 %	27932				
10.0	10.000	9.994	-0.006	37276	0.0 %	37276				

LB85912 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Feb 2017 12:47:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-10	-	0.0000	7470A	PPB	20 Feb 2017 12:49:36	S	Std
0.2 - 1	50.2	707	-	0.2000	7470A	PPB	20 Feb 2017 12:51:45	S	Std
2.5 - 1	52.5	9634	-	2.5000	7470A	PPB	20 Feb 2017 12:53:53	S	Std
5.0 - 1	55	18683	-	5.0000	7470A	PPB	20 Feb 2017 12:56:01	S	Std
7.5 - 1	57.5	27932	-	7.5000	7470A	PPB	20 Feb 2017 12:58:09	S	Std
10.0 - 1	510	37276	-	10.0000	7470A	PPB	20 Feb 2017 13:00:16	S	Std
ICV37 - 1	ICV37	15212	4.0679	-	7470A	PPB	20 Feb 2017 13:02:48	U	SMPL
ICB37 - 1	ICB37	-41	-0.0288	-	7470A	PPB	20 Feb 2017 13:04:54	U	SMPL
CCV40 - 1	CCV40	18790	5.0290	-	7470A	PPB	20 Feb 2017 13:07:00	U	SMPL
CCB40 - 1	CCB40	-46	-0.0302	-	7470A	PPB	20 Feb 2017 13:09:07	U	SMPL
CRA - 1	CRA	809	0.1995	-	7470A	PPB	20 Feb 2017 13:11:14	U	SMPL
HighStd - 1	HighStd	37862	10.1515	-	7470A	PPB	20 Feb 2017 13:13:20	U	SMPL
ChkStd - 1	ChkStd	25838	6.9220	-	7470A	PPB	20 Feb 2017 13:15:27	U	SMPL
PB97032BL - 1	PBW	-79	-0.0390	-	7470A	PPB	20 Feb 2017 13:17:34	U	SMPL
PB97032BS - 1	LCSW	14177	3.7900	-	7470A	PPB	20 Feb 2017 13:19:42	U	SMPL
I1884-13 - 1	MANHOLE-001	-5	-0.0192	-	7470A	PPB	20 Feb 2017 13:21:49	U	SMPL
I1884-13DUP - 1	MANHOLE-001DUP	1	-0.0176	-	7470A	PPB	20 Feb 2017 13:23:58	U	SMPL
I1884-13MS - 1	MANHOLE-001MS	16693	4.4657	-	7470A	PPB	20 Feb 2017 13:26:06	U	SMPL
I1884-13MSD - 1	MANHOLE-001MSD	16991	4.5458	-	7470A	PPB	20 Feb 2017 13:28:14	U	SMPL
I1884-16 - 1	MANHOLE-002	-15	-0.0219	-	7470A	PPB	20 Feb 2017 13:30:21	U	SMPL
CCV41 - 1	CCV41	18818	5.0365	-	7470A	PPB	20 Feb 2017 13:32:28	U	SMPL
CCB41 - 1	CCB41	-50	-0.0313	-	7470A	PPB	20 Feb 2017 13:34:35	U	SMPL
I1884-19 - 1	GROUNDWATER	37	-0.0079	-	7470A	PPB	20 Feb 2017 13:36:42	U	SMPL
I1884-13LX5 - 1		-13	-0.0213	-	7470A	PPB	20 Feb 2017 13:38:49	U	SMPL
I1884-13A - 1		15126	4.0448	-	7470A	PPB	20 Feb 2017 13:40:56	U	SMPL
CCV42 - 1	CCV42	18629	4.9857	-	7470A	PPB	20 Feb 2017 13:43:03	U	SMPL
CCB42 - 1	CCB42	-44	-0.0296	-	7470A	PPB	20 Feb 2017 13:45:11	U	SMPL