


245.1

LB85913
INSTRUMENT ID: CV2

Linear

 μ Abs.:

34836



A= 0.0000e+000

B= 2.8471e-004

C= -2.0528e-002

Rho= 0.9998253

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.024	-0.024	-13	0.000	-13				
0.2	0.200	0.188	-0.012	733	0.0 %	733				
2.5	2.500	2.509	0.009	8885	0.0 %	8885				
5.0	5.000	4.998	-0.002	17627	0.0 %	17627				
7.5	7.500	7.631	0.131	26876	0.0 %	26876				
10.0	10.000	9.898	-0.102	34836	0.0 %	34836				

LB85913 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Feb 2017 13:58:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-13	-	0.0000	245.1	PPB	20 Feb 2017 14:00:03	S	Std
0.2 - 1	50.2	733	-	0.2000	245.1	PPB	20 Feb 2017 14:02:11	S	Std
2.5 - 1	52.5	8885	-	2.5000	245.1	PPB	20 Feb 2017 14:04:19	S	Std
5.0 - 1	55	17627	-	5.0000	245.1	PPB	20 Feb 2017 14:06:28	S	Std
7.5 - 1	57.5	26876	-	7.5000	245.1	PPB	20 Feb 2017 14:08:35	S	Std
10.0 - 1	510	34836	-	10.0000	245.1	PPB	20 Feb 2017 14:13:15	S	Std
ICV38 - 1	ICV38	13634	3.8612	-245.1		PPB	20 Feb 2017 14:15:53	U	SMPL
ICB38 - 1	ICB38	-28	-0.0285	-245.1		PPB	20 Feb 2017 14:17:59	U	SMPL
CCV43 - 1	CCV43	17365	4.9234	-245.1		PPB	20 Feb 2017 14:20:05	U	SMPL
CCB43 - 1	CCB43	-41	-0.0322	-245.1		PPB	20 Feb 2017 14:22:12	U	SMPL
CRA - 1	CRA	715	0.1830	-245.1		PPB	20 Feb 2017 14:24:19	U	SMPL
HighStd - 1	HighStd	34864	9.9056	-245.1		PPB	20 Feb 2017 14:26:25	U	SMPL
ChkStd - 1	ChkStd	24182	6.8643	-245.1		PPB	20 Feb 2017 14:28:32	U	SMPL
PB97031BL - 1	PBW	-63	-0.0385	-245.1		PPB	20 Feb 2017 14:30:40	U	SMPL
PB97031BS - 1	LCSW	13921	3.9429	-245.1		PPB	20 Feb 2017 14:32:47	U	SMPL
I1829-01 - 1	32220008-3	-46	-0.0336	-245.1		PPB	20 Feb 2017 14:34:55	U	SMPL
I1880-01 - 1	MANHOLE	53	-0.0054	-245.1		PPB	20 Feb 2017 14:37:03	U	SMPL
I1880-01DUP - 1	MANHOLEDUP	48	-0.0069	-245.1		PPB	20 Feb 2017 14:39:11	U	SMPL
I1880-01MS - 1	MANHOLEMS	13986	3.9614	-245.1		PPB	20 Feb 2017 14:41:19	U	SMPL
I1880-01MSD - 1	MANHOLEMSD	14353	4.0659	-245.1		PPB	20 Feb 2017 14:43:26	U	SMPL
CCV44 - 1	CCV44	17202	4.8770	-245.1		PPB	20 Feb 2017 14:45:33	U	SMPL
CCB44 - 1	CCB44	-40	-0.0319	-245.1		PPB	20 Feb 2017 14:47:40	U	SMPL
I1897-01 - 1	021417-PRRT-F-D-B	114	0.0119	-245.1		PPB	20 Feb 2017 14:49:47	U	SMPL
I1897-02 - 1	021417-PRRT-F-D-S	106	0.0097	-245.1		PPB	20 Feb 2017 14:51:54	U	SMPL
I1897-03 - 1	021417-PRRT-F-U-M	180	0.0307	-245.1		PPB	20 Feb 2017 14:54:01	U	SMPL
I1904-01 - 1	021517-PRRT-F-D-B	122	0.0142	-245.1		PPB	20 Feb 2017 14:56:08	U	SMPL
I1904-02 - 1	021517-PRRT-F-D-S	80	0.0022	-245.1		PPB	20 Feb 2017 14:58:15	U	SMPL
I1904-03 - 1	021517-PRRT-F-U-B	84	0.0034	-245.1		PPB	20 Feb 2017 15:00:23	U	SMPL
I1904-04 - 1	021517-PRRT-F-U-S	75	0.0008	-245.1		PPB	20 Feb 2017 15:02:31	U	SMPL
I1906-01 - 1	021617-PRRT-E-D-M	69	-0.0009	-245.1		PPB	20 Feb 2017 15:04:39	U	SMPL
I1906-02 - 1	021617-PRRT-E-U-M	91	0.0054	-245.1		PPB	20 Feb 2017 15:06:48	U	SMPL
CCV45 - 1	CCV45	16943	4.8033	-245.1		PPB	20 Feb 2017 15:08:55	U	SMPL
CCB45 - 1	CCB45	-39	-0.0316	-245.1		PPB	20 Feb 2017 15:11:02	U	SMPL
I1880-01LX5 - 1		-9	-0.0231	-245.1		PPB	20 Feb 2017 15:13:09	U	SMPL
I1880-01A - 1		12848	3.6374	-245.1		PPB	20 Feb 2017 15:15:16	U	SMPL
CCV46 - 1	CCV46	17187	4.8728	-245.1		PPB	20 Feb 2017 15:17:22	U	SMPL
CCB46 - 1	CCB46	-36	-0.0308	-245.1		PPB	20 Feb 2017 15:19:29	U	SMPL