


7471A

LB85916
INSTRUMENT ID: CV4

Linear

 μ Abs.:

47513



A= 0.0000e+000

B= 2.1003e-004

C= 7.7611e-003

Rho= 0.9999849

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	12	0.000	12				
0.2	0.200	0.206	0.006	943	0.0 %	943				
2.5	2.500	2.461	-0.039	11680	0.0 %	11680				
5.0	5.000	5.018	0.018	23856	0.0 %	23856				
7.5	7.500	7.518	0.018	35755	0.0 %	35755				
10.0	10.000	9.987	-0.013	47513	0.0 %	47513				

LB85916 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 20 Feb 2017 15:38:39

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	S0	12	-	0.00007471A	PPB	20 Feb 2017 15:41:06	S	Std
0.2 - 1	S0.2	943	-	0.20007471A	PPB	20 Feb 2017 15:43:24	S	Std
2.5 - 1	S2.5	11680	-	2.50007471A	PPB	20 Feb 2017 15:45:41	S	Std
5.0 - 1	S5	23856	-	5.00007471A	PPB	20 Feb 2017 15:47:59	S	Std
7.5 - 1	S7.5	35755	-	7.50007471A	PPB	20 Feb 2017 15:50:16	S	Std
10.0 - 1	S10	47513	-	10.00007471A	PPB	20 Feb 2017 15:52:32	S	Std
ICV39 - 1	ICV39	18305	3.8524	-7471A	PPB	20 Feb 2017 15:55:15	U	SMPL
ICB39 - 1	ICB39	-23	0.0029	-7471A	PPB	20 Feb 2017 15:57:31	U	SMPL
CCV47 - 1	CCV47	23634	4.9717	-7471A	PPB	20 Feb 2017 15:59:46	U	SMPL
CCB47 - 1	CCB47	-29	0.0017	-7471A	PPB	20 Feb 2017 16:02:02	U	SMPL
CRA - 1	CRA	945	0.2062	-7471A	PPB	20 Feb 2017 16:04:18	U	SMPL
HighStd - 1	HighStd	47482	9.9806	-7471A	PPB	20 Feb 2017 16:06:34	U	SMPL
ChkStd - 1	ChkStd	31622	6.6494	-7471A	PPB	20 Feb 2017 16:08:50	U	SMPL
PB97033BL - 1	PBS	-59	-0.0046	-7471A	PPB	20 Feb 2017 16:11:06	U	SMPL
PB97033BS - 1	LCSS	16122	3.3939	-7471A	PPB	20 Feb 2017 16:13:23	U	SMPL
I1884-01 - 1	PIT-A-COMP	12788	2.6937	-7471A	PPB	20 Feb 2017 16:15:40	U	SMPL
I1884-04 - 1	PIT-B-COMP	27044	5.6879	-7471A	PPB	20 Feb 2017 16:17:58	U	SMPL
I1884-07 - 1	PIT-C-COMP	9396	1.9812	-7471A	PPB	20 Feb 2017 16:20:15	U	SMPL
I1884-10 - 1	PIT-D-COMP	13932	2.9340	-7471A	PPB	20 Feb 2017 16:22:33	U	SMPL
I1890-02 - 1	24B-7-9	140	0.0372	-7471A	PPB	20 Feb 2017 16:24:49	U	SMPL
CCV48 - 1	CCV48	23931	5.0341	-7471A	PPB	20 Feb 2017 16:27:05	U	SMPL
CCB48 - 1	CCB48	-32	0.0010	-7471A	PPB	20 Feb 2017 16:29:21	U	SMPL
I1890-02DUP - 1	24B-7-9DUP	205	0.0508	-7471A	PPB	20 Feb 2017 16:31:37	U	SMPL
I1890-02MS - 1	24B-7-9MS	16342	3.4401	-7471A	PPB	20 Feb 2017 16:33:53	U	SMPL
I1890-02MSD - 1	24B-7-9MSD	16133	3.3962	-7471A	PPB	20 Feb 2017 16:36:08	U	SMPL
I1892-05 - 1	TP22-COMP16	50710	10.6586	-7471A	PPB	20 Feb 2017 16:38:24	U	SMPL
I1899-01 - 1	SS050SS219(0-6)-170216	29878	6.2831	-7471A	PPB	20 Feb 2017 16:40:41	U	SMPL
I1913-01 - 1	SP-ABCD-COMP	5918	1.2507	-7471A	PPB	20 Feb 2017 16:42:58	U	SMPL
I1913-06 - 1	SP-EFGH-COMP	3023	0.6427	-7471A	PPB	20 Feb 2017 16:45:15	U	SMPL
I1890-02LX5 - 1		29	0.0139	-7471A	PPB	20 Feb 2017 16:47:32	U	SMPL
I1890-02A - 1		18304	3.8522	-7471A	PPB	20 Feb 2017 16:49:50	U	SMPL
I1892-05DLX2 - 1		25483	5.3601	-7471A	PPB	20 Feb 2017 16:52:06	U	SMPL
CCV49 - 1	CCV49	23885	5.0244	-7471A	PPB	20 Feb 2017 16:54:22	U	SMPL
CCB49 - 1	CCB49	-19	0.0038	-7471A	PPB	20 Feb 2017 16:56:39	U	SMPL